

Amendment –VIII

Dated 13.11.2018

Tender Ref : HSCC/Mauritius/ENT hospital Equipment/2018 dated 22.09.2018

Sub: Procurement of Medical Equipment for New ENT Hospital Mauritius

Ref.: Tender No. HSCC/Mauritius/ENT Hospital Eqpt./2018 dated 22.9.2018.

Technical Amendments are as follows:

Item No. 9 Neuro Navigation System

Sr. No.	Tender Specification	To be Read As
	System should offer hybrid optical technology Wireless Active or Passive.	System should offer optical technology or Electromagnetic Technology.
	System should be supplied with Zeiss Pentero Microscope Intergration.	Deleted
	Spine Software	Deleted

Item No. 8 Nerve Stimulator Locater and Mapper

Sr. No.	Tender Specification	To be Read As
	Nerve stimulation mode should have Twitch, TOF, Tetanus, Double Burst stimulation and Post Tetanic Count.	The System should be capable of monitoring cranial Motor Nerves III, IV, V, VI, VII, IX, X, XI and XII.

Item No.13 Shaver System Cum Micro Drill

Sr. No.	Tender Specification	To be Read As
	EMD : Rs.24,000/-	EMD : Rs.60,000/-

The revised bid submission date remains same i.e 19.11.2018

All other terms and conditions of the tender enquiry of tender document & and submission date shall remain unchanged.

Prospective bidders are advised to regularly visit website www.tenderwizard.com/HSCC&www.hscltd.com for corrigendum/amendments etc. if any, as there will be no separate advertisement for them.

Senior Chief Executive
Ministry of Health & Quality of Life
Govt. of Republic Mauritius , Mauritius