

**HSCC (I) Ltd**

**Amendment –VIII Dated 20.10.2020**

**Ref. IFB No: HSCC/AIIMS-RAEBARELI/Hospital/06/2020 Dated : 18.08.2020**

**Subject: [Procurement of Medical Equipment for AIIMS -Raebareli.](#)**

**Technical Amendments and Extension of Bid submission date :-**

The reply to the pre bid queries for **Eye Department & Microbiology Department** is given below:

**Sr. No. 41 Noncontact Tonometer with Pachymeter ..**

<b>SI No</b>	<b>Technical specifications</b>	<b>Decision of Technical Experts</b>
10	Motorised Table of same make for NCT	“Motorised Table suitable for NCT ”

**Sr. No. 42 ND : YAG Laser-1064nm**

No Change

**Sr. No. 43 Ophthalmic Operating Microscope**

No Change

**Sr. No. 44 Digital Funds Camera**

No Change

**Sr. No. 45 Auto Refractometer**

No Change

**Sr. No. 46 Visual Field Analyzer**

No Change

**Sr. No. 47 Ultrasound (A+B Scan)**

No Change

**Sr. No. 48 Indirect Ophthalmoscope (Wireless)**

No Change

**Sr. No. 49 High Speed Steam Flash Sterilizer**

No Change

**Sr. No. 50 Digital lensometer**

No Change

**Sr. No. 51 OT Table Eletro –Hydraulic**

No Change

**Sr. No. 52 Teller Acuity Charts**

No Change

**Sr. No. 53 Swept Source OCT with integrated OCT Angio and autofluorescence**

No Change

**Sr. No. 54 Ophthalmic Operating Microscope with Motorized Wide Angle Viewing System**

No Change

**Sr. No. 55 Vitrectomy Machine with Cut Rate Up to 10,000 and IOP Control with Inbuilt Green Laser Delivery System**

SI No	Technical specifications	Decision of Technical Experts
10	Phaco Mode: 2)Should have the facility to use variety of phaco tips like Kelman, ABS and micro tips	Should have the facility to use variety of phaco tips
12	Other Features: 5), 6), 7)	Point Deleted
13	Consumables: C)Chandelier illuminating system, Finace loop membrane scraper	Consumables: Chandelier illuminating system”

**Sr. No. 56 Phaco Machine (Torsional Technology Emulsification System)**

SI No	Technical specifications	Decision of Technical Experts
8.	The Torsional or non - Longitudinal or Transversal ultrasound blended handpiece should be available to drive latest generation Phaco Tip Like Kelmen/Angled Tip, Flared, Min Flared And/Or Aspiration Bypass ( ABS ) Tips in Both1.1 Or 0.9 mm Configuration	The Torsional or non - Longitudinal or Transversal ultrasound blended handpiece should be available to drive latest generation Phaco Tip Like Kelmen/Angled Tip, Flared <b>Or Mini</b> Flared And/Or Aspiration Bypass ( ABS ) Tips in Both1.1 Or 0.9 mm Configuration
12.	Equipped with option 1.8 mm, 2.2 mm , 2.8 mm of Phaco Tip	Equipped with option <b>1.8 mm or 2.2 mm</b> , 2.8 mm of Phaco Tip

**Sr. No. 57 Non-Contact Optical Coherence Biometer**

SI No	Technical specifications	Decision of Technical Experts
1	Should be based on most advanced SS OCT based technology	Should be based on most advanced SS OCT based or Optical Low Coherence Reflectometry based(OLCR) ”
2	Axial Length 14-38mm	Axial Length 14-32mm or more”
4	Lens Thickness 1-10mm(Phakic Eye)	Lens Thickness 0.5-6.5mm or more
6	Central Corneal Thickness 0.2-1.2mm	Central Corneal Thickness 0.3-0.8mm or more”

**Sr. No. 58 Ultrasound Biomicroscope (UBM Probe 35 MHz 50 MHz)**

No Change

**Item No. 10 : Real Time PCR**  
**Department : Microbiology**

<b>SI No</b>	<b>Technical specifications</b>	<b>Decision of Technical Experts</b>
Point 10	Temperature Minimum Range 4-100 °C	Temperature Minimum Range 25 °C -99 °C

The Bid submission date of all above items extended from **20.10.2020 to 27.10.2020.**

The Bid submission date is extended form 20.10.2020 to 27.10.2020 of equipment under **Forensic Department** for **item No.34**, item no. 33, & 35.**Microbiology** item No. 4,5,7,8,9, for **Urology items** 36,37,39, and **Pharmacology** Item No.15. for **Lab Medicine** Item No. 18,19,20,25,27, 28, and 29 .

In view of the above, the bid submission date for all above items is extended from **20.10.2020 to 27.10.2020.**

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website,CPPP, Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

**CGM (PROC.)**  
**HSCC (I) LTD.**