

Amendment-IX

Ref.: Tender Enquiry No. HSCC/PUR/LHMC/2014 dt. 16.7.2014.

Sub.: Procurement of Radiotherapy Equipment for Lady Hardinge Medical College & Hospital,
New Delhi.

Amendment –IX (For Schedule 1- High Energy Linear Accelerator)

Existing Specifications:

8. Image-Guided Radiotherapy System

- 8.1 Kilovoltage-based 3D-Image-Guided Radiotherapy (kV-IGRT) shall be provided and it should have FDA clearance. The system shall have the capability of producing 2D radiography, 2D fluoroscopy and 3D (volume) cone beam CT imaging modalities for patient's interfraction and intrafraction daily setup verification.

Amended Specifications:

8. Image-Guided Radiotherapy System

- 8.1 Kilovoltage-based 3D-Image-Guided Radiotherapy (kV-IGRT) shall be provided and it should have FDA clearance. The system shall have the capability of producing 2D radiography, 2D fluoroscopy and 3D (volume) cone beam CT imaging modalities for patient's interfraction daily setup verification.

Rest all tender conditions and Amendments issued so far remain unchanged.

**Director,
LHMC, New Delhi**