

All Bidders

Amendment –IX

Project: Supply, Installation, Testing & Commissioning of CSSD for Super Specialty Block at Indira Gandhi Medical College, Shimla, H.P.

IFB No. HSCC/SES/CSSD/SSB/Shimla/2019 dated 13.12.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender Document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Amended as
Page No.7, Item No.4 Point No.4d	<p>Washer disinfectant (300 to 350L) with accessories</p> <p>Chamber Capacity: Chamber capacity: Operational Volume should be 300 to 350 L. Should supply 12 Nos of standard Stainless Steel DIN trays. It should be able to process minimum 12 DIN trays (Approx480x250x50) with 6 levels in single process.</p> <p>Amendment required :</p> <p>12 DIN Trays are Company Specific. You are requested to kindly amend it to 10-12 DIN Trays.</p>	<p>Double Door washer disinfectant 250-300 litre (10-15 Din Trays) with accessories. Should supply 15 nos 10-15 DIN trays.</p>
Page-7, Point No.4	<p>The unit should also have an interface as standard for an optional batch printer.</p> <p>Amendment required :</p> <p>Please modify it to unit should have an interface with RS232.</p>	<p>RS232 port connectivity is allowed</p>
Page No.B, Item No.6 Point No.6a	<p>Ultrasonic Cleaner (40 L)</p> <p>The units should be a compact bench top model, with a built-in tank manufactured from high quality AISI(316/304) stainless steel and a solid-state generator that sends Ultrasonic (Approx 40 KHz)</p>	<p>The units should be a compact bench top model, with a built-in tank manufactured from high quality AISI(316/304) stainless steel and a solid-state generator that sends Ultrasonic (25- 40 KHz)</p>

	<p>Amendment required</p> <p>The frequency of 40 KHz may be Company specific. Kindly amend the same to 25 - 40 KHz</p>	
Item No. 6e	<p>It should have digital read out timer and temperature setting (temperature adjustable from 30 -69°C or more) monitoring.</p> <p>Amendment required</p> <p>Kindly amend temperature adjustable setting to 25 – 75 deg C</p>	<p>It should have digital read out timer and temperature setting (temperature adjustable from 25 - 75°C or more) monitoring.</p>
Page No.1 Item No.-1 b)-1	<p>Horizontal Sterilizer 550 Ltr-600L or more with Accessories</p> <p>The Chamber and doors should be made of solid, high quality 316L Stainless steel.</p> <p>Amendment required</p> <p>Kindly amend chamber and doors shall be made of SS 304 / SS 316 L Stainless steel.</p>	<p>The chamber and doors should be made of solid, high quality AISI 316L Stainless Steel</p>
Page No.3, Item No.2 Page No.4, Point No.(B)- 1	<p>HIGH SPEED STERILIZER Double door 250 L wth Accessories:</p> <p>The chamber and doors should be made of solid, high quality 316L Stainless steel.</p> <p>Amendment required</p> <p>Kindly amend chamber and doors shall be made of SS 304 /SS 316 L Stainless steel</p>	<p>The chamber and doors should be made of solid, high quality AISI 316L Stainless Steel</p>

Page No.5, Point No.(e)-2	These access levels should be user selectable. The control system should have CPU processor with battery back-up & non-volatile memories, Digital input/output controls, analogue measuring inputs & COM ports for Printer & PC connectivity. Amendment required Please allow RS232 port connectivity	RS232 port connectivity is allowed
---------------------------------	--	------------------------------------

The bid submission date is extended from 03.02.2020 to 06.02.2020 and bid security should be valid for 180 days from the date of original bid submission ie. from 03.01.2020.

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.
For and behalf of Principal, IGMC, Shimla