

All Bidders**Amendment-IX**

Subject: Supply, Installation, Testing & Commissioning of Medical Gas Pipeline System for Lady Hardinge Medical College, New Delhi.

Tender No: HSCC/SES/MGPS/LHMC/2021 Date: 22.12.2021

This has reference to above tender No.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Reply
1.	<p>Page no. 19 Vol04_TechSpecs 12. Alarm System 12.2 Medical Gas Area Alarm</p> <p>As per Amendment-VI dated 18.02.2022</p> <p>Alarm System should be Touch Type Alarm Technology (7 inch or more) On screen key pads allow for simple editing of text fields.</p> <p>Sir, we once again wish to re-iterate that worldwide different manufacturers produce different sizes of alarm. So it is not necessary that all the manufacturers will provide 6.5 inches or more screen.</p> <p>We request screen size should be 2.85 inch or more which will cover all the manufacturer's made screen sizes and with wide range more bidders would be able to participate.</p> <p>We request to kindly amend the size of screen/display to 2.85" or above to neutralize the size for all bidders.</p>	<ul style="list-style-type: none"> • Single screen alarm • Good visibility from distance • Good visibility of all the required services as per the areas
2.	<p>COMMON</p> <p>Free Sale certificate notarized by the chamber of commerce must be submitted.</p>	Deleted.

	<p>Sir kindly understand free sale certificate is specifically required for exporting free selling of Drug & pharmaceutical and that too on certain Drugs who's quantities are in short supply.</p> <p>Whereas, Medical Gas Pipeline Equipments DO NOT require any Free Sale Certificate from notarized chamber of commerce because it is already free selling/export equipment. Neither it is prohibited nor restricted for sale. NO Restrictions for MGPS items is applied by any country to sell it to our India.</p> <p>It is request that this clause should be deleted as this is unnecessary, manipulative clause to deviate the bidders and Department. Sir the second proof is we have been contractor of M/s HSCC (I) Ltd. and other Govt. agencies from last more than 10 years, without any restrictions we are importing & supplying MGPS products. Then why this certificate unnecessary nor does M/s HSCC had asked in any of the earlier tenders.</p>	
3.	<p>COMMON</p> <p>Request for Extension of Bid Submission Date</p> <p>Sir, in view of recent Amendment-VI dated 18.02.2022 to the Tender, we request you to kindly extend the Bid Submission date by atleast 1 week in order to take cognizance of amendments made and our above request for further amendments.</p>	<p>Bid submission date has already been extended till 09.03.2022.</p>
4.	<p>As per Amendment No. VI</p> <p>Alarm System should be touch type Alarm technology (6.5 inches or more). On screen key pad allow for simple editing of text fields including main key board, alternative key board & number pad.</p> <p>Request for Amendment We request you to kindly amend, Area Alarm touch type screen to 5.5 inches or more instead of 6.5 inches or more as our OEM provides the same. More over this does not provide any functional advantage.</p> <p>Alarm system should be touch type Alarm technology with inbuilt system. Additional</p>	<ul style="list-style-type: none"> • Single screen alarm • Good visibility from distance • Good visibility of all the required services as per the areas

	attachments like on screen key pads, alternative key board and number pads are not required.	
5.	Amendment No. III Page No. 36, point no. 93, Free Sale Certificate This is specific to few European Companies. We represent American Company where the same is not applicable however our letter of Authorisation says that we are the authorized vendor and our Principal Company is committed to provide spare parts till the life time of the product.	Deleted.
6.	Bid Submission date extension We request you to kindly extend the due date by 07-10 working days to enable wider participation, as it is elaborate tender & lot of documents have to be arranged from our Manufacturers to negotiate the best possible pricing and terms.	Bid submission date has already been extended till 09.03.2022.
7.	OxygenFlowmeter withhumidifierbottle, Ward VacuumUnit, Theatre Suction Unit, Isolation valve, AlarmSystem, BedHead Panels&AreavalveServiceSystem. Proposed Amendment: The Werequestyouto, sincethese products areavailable indigenously, henceasper Government notification forpreference for 'Make inIndia'products, &youhavetoshow your interest too &budgetisverylowforall items,while youhaveaskedforallthe imported.So,werequestyoutokindly consider these productsasIndian &youwill nothaveto increasethe budgetofthe Tenderalso.Weare installed theseproducts(verygoodQuality) in allIPMSSYHospital alloverIndia.	Originaltender specifications will prevail for these items: OxygenFlowmeter withhumidifier bottle, Ward VacuumUnit, Theatre Suction Unit, Isolation Valve, BedHead Panels&AreavalveServiceSystem.

All other terms & conditions remain unchanged.

Director, LHMC, New Delhi