

HSCC (I) Ltd

Amendment –IX Dated 28.01.2020

Tender Ref : HSCC/AIIMS-RAEBARELI/Hospital/2/2019 Dated : 28.11.2019

Department: Others Equipment

Item No. 2: Syringe Infusion Pump, Qty -155 Nos

Spec. Sr. No.	TENDER SPECIFICATION	READ AS
Pt. No. 15	<u>Mounting Device/ Docking Station</u> for atleast four pumps as per requirement so as to enable to power up to 4 pumps with one power cord when mounted on IV pole (Price to be quoted separately.)	<u>Docking Station</u> for at least four pumps as per requirement so as to enable to power up to 4 pumps with one power cord when mounted on IV pole (Price to be quoted separately.)
Pt.No.5	Bolus rate programable from <u>40 to 1000 ml/hr.</u> or more with infused volume display and one key press bolus. Reminder audio after every 1 ml delivered /programmable bolus should be available	Bolus rate programable from <u>40 to 1000 ml/hr.</u> or more with infused volume display and one key press bolus..
Pt. No. 11	Manual/automatic pusher	Manual/automatic pusher engagement after loading of syringe
Pt. No. 13	Should have comprehensive alarm package including Occlusion limit exceed alarm, Near end.	Should have comprehensive alarm package including Occlusion limit exceed alarm, Near end, KVO, Low Battery.
BOQ	2. Docking Station 1No	Docking Station 10 Nos.

Item No.9, Defibrillator with ECG monitor, Qty-15 Nos

Spec. Sr. No.	TENDER SPECIFICATION	READ AS
Pt. No. 7.3	Should have local service facility.	Equipment compliant will be attended in within 24 hrs. of log in the complaint to the service provider.
	Point Added	Should be US-FDA or European CE with 4 Digit notified body number or BIS

Item No.10, Bi PAP/CPAP

Spec. Sr. No.	TENDER SPECIFICATION	READ AS
	Point Added	Should be US-FDA or European CE with 4 Digit notified body number . or BIS

Item No.11: ECG

Spec. Sr. No.	TENDER SPECIFICATION	READ AS
Point 1	Twelve Channel 5.7” or more LCD display for all 12 leads along with on screen details.	Twelve Channel 7” or more LCD display for all 12 leads along with on screen details.
	New Point Added	US-FDA or European CE with 4 Digit notified body number . or BIS

7	Waveforms can be recorded	Waveforms can be recalled & printed again
	Point Added	system should have LAN and FTP facility

Item No.12: Pulse Oximeter

Spec. Sr. No.	TENDER SPECIFICATION	READ AS
	New Point Added :	Should be US-FDA or European CE with 4 Digit notified body number . or BIS

In view of the above amendments the revised submission date of Eppt Sr. No. N2, N09, N10, N11 & N12 is now **07.02.2020**

Tender of Radiology Equipment (**128 –Slice CT Scan, 1.5 Tesla MRI, Digital X-Ray 1000 m.A System, and Digital Radiofluoroscropy System**) **Stand Cancelled**. Amendments of other Radiology Equipment (Mobile X Ray, Color Doppler (2D & 3 D) , Portable Color Doppler System, CR System and Mammography System) are pending. Meantime their bid submission date is extended to **07.02.2020**

Amendments of Equipments under Orthopedics Department, CTVS Department, Neurosurgery Department Surgery Department , ENT Department , Gyne & Obst. Department , Blood Bank and PMR Department and Item No. N8 (C-Arm with IITV) and Item No. N6C are pending. Meantime, their bid submission date is also extended to **07.02.2020**

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPPP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard

**For CGM (P)
HSCC (I) Ltd**