

HSCC (I) Ltd

Amendment –IX Dated 01.02.2021

Ref. IFB No: HSCC/AIIMS-RAEBARELI/Hospital/08/2020 Dated :20.11.2020

Subject: Procurement of Medical Equipment for AIIMS -Raebareli.

Technical Amendment:

Item No. 44 Name of Item **Refrigerated Blood Bag Centrifuge (12 Bags)**

Department: Blood Bank

Specification Sr. No.	Tendered Specification	To be Read As
iii(a)	Minimum Speed 5000 RPM & above.	Minimum Speed 5000G & above.
iii-b:-	RCF(Relative Centrifugal Force)for blood bags: 6000-6500 G	Point III a and III b should be changed to RCF (Relative Centrifugal force) it should be able to deliver more than 6500 with full laod of blood bags.
Point F 2	Twelve 350 ml and or 450 ml single, double, triple, quadruple/ quintuple, blood bags with SAGM bag and empty satellite bags with inline filter system.	It should accommodate sixteen 350 ml and / or 450 ml single, double, triple , quadruple/quintuple Blood Bags with SAGM bag and empty satellite bags and inline filter.

Item No. 54 Name of Item **Platelet Agitator cum Incubator,(90 Bags)**

Department: Blood Bank

Specification Sr. No.	Tendered Specification	To be Read As
1	Platelet incubator should have the provision to store 96 platelet bags agitator	Platelet incubator should have the provision to store 96 or more platelet bags agitator.

All other specifications shall remain unchanged.

Item No. 48 Name of Item , **Mobile Transport Box**

Department: Blood Bank

Specification Sr. No.	Tendered Specification	To be Read As
2	Should be robust, light weight, portable mobile refrigerated transport box made up of rotationally moulded polyurethane	Should be robust, light weight, portable mobile refrigerated transport box made up of rotationally moulded polyurethane/ Powder Coated CRCA Steel.

All other specifications shall remain unchanged.

Item No. 50 Name of Item ,**Table Top Micro plate Centrifuge**

Department: Blood Bank

Specification Sr. No.	Tendered Specification	To be Read As
04	4:- Manufacturer should provide compatible set of microplates with specifications of the same	Deleted
25	25:- whenever there is breakdown the firm will carry out the repair within 48 hours of receipt of such Information(either by telephone or by other means)	Whenever there is breakdown the firm will attend the breakdown call within 48 hours of receipt of such Information (either by telephone or by other means) & close it in next 5-7 working days.
26	26:- If there is delay beyond 48 hours than the firm will be penalized at the rate of 1% of the cost of product per day.	No Change

All other specifications shall remain unchanged.

Item No. 51 Name of Item , **Blood Bank Refrigerator**
Department: Blood Bank

Specification Sr. No.	Tendered Specification	To be Read As
2.7	Insulation of high grade pressure foam material greater than 80 mm thick CFC Free	Insulation should be of high density PUF with thickness of 70- 80 mm. CFC Free.

All other specifications shall remain unchanged.

Item No. 49 Name of Item : **Aphaeresis Machine**
Department: Blood Bank

Specification Sr. No.	Tendered Specification	To be Read As
1	Continuous flow blood cell Separator	Intermittent/ Continuous flow blood cell Separator
6	Built in Leukoreduction (5x10 ⁶) for platelets & plasma using elutriation (e.g LRS chamber) or other patented technology which is not based on leukoreduction filter.	Built in Leucoreduction for platelets and plasma which is not based on leukoreduction filter.
Point 13 a	New point added	Minimum Inlet flow rate should be 5 ml/ min

Item No, 41 Name of Item : **Biosealer (4 No. for Blood Bank)**

All specifications shall remain unchanged.

Item No. 42 Name of Item : **Biosealer (Hand Held) (2 No. for Blood Bank)**

All specifications shall remain unchanged.

Item No. 43 Name of Item : **Folding Donor Couch (One Set -2 chair with one trolley (2 No. for Blood Bank)**

All specifications shall remain unchanged.

Item No. 46 Name of Item : **Sterile connecting Device (1 No. for Blood Bank)**

All specifications shall remain unchanged.

Item No. 47 Name of Item : **Gel/ Bead Centrifuge and Incubator (2 No. for Blood Bank)**

All specifications shall remain unchanged.

Item No. 52 Name of Item : **Donor Couch (4 No. for Blood Bank)**

All specifications shall remain unchanged.

Item No. 55 Name of Item : **Laminar Flow (Small) (1 No. for Blood Bank)**

All specifications shall remain unchanged.

Item No. 45 Name of Item: -40 ° C Deep Freezer

All specifications shall remain unchanged.

All other specifications shall remain unchanged All other items of Blood Bank specifications shall remain unchanged and the bid submission date remains 03.02.2021 and bid opening date to 04.02.2021

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPPP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

**CGM (PROC.)
HSCC (I) LTD**