

HSCC (I) Ltd

Amendment –IX Dated 12.11.2020

Ref. IFB No: HSCC/SJH/Medical Equipment/2020/45 Dated: 21.08.2020

The Technical amendments are as follows:

Name of Item S.No.Item No. 1 Adult Cardiac Instrument Set.

Specification Sr. No.	Tendered Specification	To be Read As
Adult Cardiac Instrument set - Point S.no. 34	Debakey Needle Holder 8" TC 23cm	Debakey Needle Holder 8 to 8.25 inch TC
Adult Cardiac Instrument set - Point S.no. 35	9" TC 26cm	Debakey Needle Holder 9 inch TC
Adult Cardiac Instrument set - Point S.no. 101	Hegar Dilators 1mm to 28mm (Set of 14 Nos)	Hegar Dilators 1mm to 28mm if one size per dilator [set of 28] if 2 sizes per dilator [set of 14]
	New Point Added	Adult Cardiac instrument set - Conatiner Size 23" x 11" x 10" (2qty) and Perforated basket with Lid 23" x 11" x 4" (Qty 4) per set.
BOQ	To Qualify, Bidders must quote at least 70% of the tendered quantity. This supercedes the qualification criteria of the tender documentation	To Qualify, Bidders must quote at least 70% of the tendered quantity from a single manufacturer
Company should have service facility in India.	We request to change this as " Company or its authorised Dealer should have service facility in India	Company or its authorized Dealer should have service facility in India
New point addition in BOQ	+ - 10% variation in dimension of the instrument is acceptable.	+ - 10% variation in dimension of the instrument is acceptable

Name of Item S.No.Item No. 2 Pediatric Cardiac Instrument Set.

Specification Sr. No.	Tendered Specification	To be Read As
Pediatric Instrument Set Point No. 29	New born Finochitto Rib Retractor Single Blade W 12 D14 mm arm spread 54 mm arm length 50 mm	Read as "Finochitto Rib Retractor Single Blade W 15 D 15 spread 80 mm arm length 56 mm"
Pediatric Instrument Set - Point Sr. no - 30	Finochitto Rib Retractor Single Blade New Born D 6 x W 7, spread 37 mm, arm length 25cm	Read as "Finochitto rib retractor new born D 7 mm W 7 mm spread 36 mm , Arm length 37 mm"
Pediatric Micro Instrument set - Point number 10-11	Coronary Spring Potts Scissor Round / flat Handle 45 Degree, 7" , 11 mm blade and Coronary Spring Potts Scissor Round / flat Handle 125 Degee, 7" 11 mm Blade	Coronary Spring Potts Scissor Round / flat Handle 45 Degree, 7" , 9-11 mm blade and Coronary Spring Potts Scissor Round / flat Handle 125 Degee, 7" , 9-11 mm Blade
Point 20	Micro Instrument tray with silicon mat (Imported)	Pediatric instrument set [under micro instrument set] 10±0.5 inch X 10±0.5 inch X 1.5 inch Aluminum with silicone inserts
New Point Added	New Point Added	2. Pediatric instrument Set - Conatiner Size 23" x 11" x 10" (2qty) and Perforated basket with Lid 23" x 11" x 4" (Qty 4) per set.
BOQ	To Qualify, Bidders must quote at least 70% of the tendered quantity. This supercedes the qualification criteria of the tender	To Qualify, Bidders must quote at least 70% of the tendered quantity from a single manufacturer

	documentation	
BOQ	Company should have service facility in India.	Company or its authorized Dealer should have service facility in India
	New point addition in BOQ	+ /- 10% variation in dimension of the instrument is acceptable

Name of Item S.No. Item No. 3 Thoracic Instrument Set

Specification Sr. No.	Tendered Specification	To be Read As
New Point Added	New Point Added	Thoracic Set - Conatiner Size 18" x 11" x 6" (2 qty) and Perforated basket with Lid 18" x 11" x 2" (Qty 4) per set
BOQ	To Qualify, Bidders must quote at least 70% of the tendered quantity. This supercedes the qualification criteria of the tender documentation	To Qualify, Bidders must quote at least 70% of the tendered quantity from a single manufacturer
BOQ	Company should have service facility in India.	Company or its authorized Dealer should have service facility in India
BOQ	New point addition in BOQ	+ /- 10% variation in dimension of the instrument is acceptable

Name of Item S.No. Item No. 4 Dual Head Gamma Camera

Sr. No.	Tendered Specification	To Read As
First line of Original Tendered Specification	16-Slice SPECT/CT Scanner - System offered to have USFDA and European CE approval certificate or BIS Standard approved.	Deleted

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website, CPPP Website & HSCC tender Wizard for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

For Medical Superintendent
VMMC & Safdarjung Hospital