

All Bidders**Amendment -X**

Subject: Supply , Installation, Testing & Commissioning of Mortuary for Chittaranjan National Cancer Institute (CNCI), Kolkata.

IFB No. : HSCC/SES/Mortuary/CNCI/Kolkata/2018 dated 28.05.2018

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sl.No	As per technical specification	Amendment
1	Working Temperature +2 °C to +8° C and humidity control. Digital temperature indication. Temperature chart recorder shall be supplied and installed.	Working Temperature +2 °C to +8° C and humidity should be maintained at 75% to 80% control. Digital temperature indication. Temperature chart recorder/USB type data logger shall be supplied and installed.
2	Loading Trolley- The carriage consists of three piece assembly which consist of stationary frame a lower carriage and an upper carriage. The lower and upper carriage assembly shall ride on wheels and tracks that allow easy telescopic action. The complete assembly is locked when returned to closed position.	Deleted (Revised BOQ is attached)
3	Turnkey works Drain system shall be inbuilt in the unit. The unit shall be supplied with suitable MCB, Wire & cables, earthing facility, suitable UPS, heavy duty exhaust fan etc. and the complete installation including other accessories and minor civil works(Construction of platform etc will be the responsibility of bidder.	Drain line work – Deleted The unit shall be supplied with Wire & cables, earthing facility, suitable UPS etc. and the complete installation. Remaining turnkey works are deleted.

	Remaining turnkey works are deleted.	
4	UPS with constant voltage supply and 30 minutes power back up.	UPS with constant voltage supply and power back up shall be supplied.
5.	LCD/TFT display.	LED/LCD/TFT Display
6.	Interior fluorescent lighting.	Interior fluorescent/ CFL/LED Lighting
7.	Swivel locking castors	Deleted
8.	The Mortuary Chamber shall be suitable for keeping 6 dead bodies and shall be complete with refrigeration system, locking arrangement, Foul Order Treatment by means of Activated Carbon filter for De-Odorizing System.	The Mortuary Chamber shall be suitable for keeping 6 dead bodies and shall be complete with refrigeration system. Drier for Refrigerant shall be supplied (For Moisture Absorption from refrigerant).

The revised BOQ is attached.

The bid submission date is extended from 24.08.2018 to 31.08.2018 and bid security should be valid for 180 days from the date of original bid submission ie. from 22.06.2018.

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.
For and behalf of Director, CNCI, Kolkata

BOQ OF MORTUARY						
Name of Bidder M/s:						
Package - Supply, Installation, Testing & Commissioning of MORTUARY on Turnkey basis with 5 years warranty						
PART-I						
Item No. 1	Description 2	Unit 3	Qty 4	Rate In Rs (in Figure) 5	Rate in Words in Rupees 6	Amount (Rs.) (In Figure) 7
1.0	Mortuary (6-Body) including UPS Complete with all accessories as per technical specification	Nos	2			0.00
2.0	TURNKEY WORKS	Lot	1			0.00
Total Rs.						0.00
Part-II						
1	Comprehensive Maintenance Charges for the completeMortuary Equipment and System including spares, repair or replacement of defective equipments/parts, tolls, tackles, accessories, consumables, labour charges etc. complete in all respect as per the contract.					
	6th Year	1	Job			0.00
	7th Year	1	Job			0.00
	8th Year	1	Job			0.00
	9th Year	1	Job			0.00
	10th Year	1	Job			0.00
SUB TOTAL Rs.						0.00
SUMMARY OF RATES QUOTED						
TOTAL (PART-I)				0.00		0.00
TOTAL (PART-II)				0.00		0.00
Grand Total Amount (PART - I + PART - II) (in Figures) :-						0.00
Grand Total in Words: Rupees						