

Amendment no. X Dated 27.09.2018

## HSCC/PUR/CNCI/Kolkata/Medical Equipment/02/A dt. 13.07.2018

Procurement of Medical Equipment CNCI 2nd Campus

All bidders are requested to note the following:

### Item no. 01- High-Energy Linear Accelerator With IGRT/IMRT/SRS/SRT Facility and Site-Modification

Tender Specifications	May Please read as
<p><b>Point V, Page No 67:</b> RADIOTHERAPY DOSIMETRY EQUIPMENT (One set for Two nos. of Linear Accelerator)</p>	<p><b>RADIOTHERAPY DOSIMETRY EQUIPMENT</b> as mentioned in the tender technical specification is common for 02 units of High Energy Linear Accelerator. Therefore one set of <b>RADIOTHERAPY DOSIMETRY EQUIPMENT</b> is to be supplied with 02 units of High Energy Linear Accelerator and quote their price accordingly.</p>
<p><b>Point VI : Page No 73 : Mould Room and Patient Fixation and Immobilization Devices/Accessories</b></p>	<p><b>Mould Room and Patient Fixation and Immobilization Devices/Accessories</b> as mentioned in the tender technical specification is common for 02 units High Energy Linear Accelerator. Therefore one set <b>Mould Room and Patient Fixation and Immobilization Devices/Accessories</b> as mentioned in tender specification is to be supplied with 02 nos. High Energy Linear Accelerator and quote their price accordingly.</p>
<p>3. Vendor will have to quote Unit Rates of the following components of Site Modification work :</p> <p>3.1 Electrical work 3.2 Air conditioning (HVAC) 3.3 Flooring 3.4 Wall Finishing &amp; Painting 3.5 False Ceiling. 5. Bidder should clearly mention break up price of each component of Site Modification work separately.</p>	<p>Bidder quote their single turnkey price including Electrical work, Air Condition (HAVC), Flooring, Wall Finishing &amp; Painting, False Ceiling, etc. or scope of work as mentioned in the tender document.</p>
<p>The system shall provide the following functions: Record and Review Patient Diagnoses; capable of recording the diagnosis as per the ICD C and ICD 10 system and complete ICD C and ICD 10 codes should be available in the system without requiring extra input, Plan a course of treatment in advance so that treatments are readily delivered when the patient arrives; Write RT prescriptions that detail treatment techniques, fractions, and dose; Define treatment fields; Link setup fields and notes to treatment fields; Setup notes should include photos that show how to set up the patient; Track dose to specific sites; Define site breakpoints with instructions that appear when the breakpoint will be exceeded; Store treatment plan information to avoid redundant and time-consuming data entry. <b>Vendor should provide the each unit price of OIS workstation offered.</b></p>	<p>The system shall provide the following functions: Record and Review Patient Diagnoses; capable of recording the diagnosis as per the ICD C and ICD 10 system and complete ICD C and ICD 10 codes should be available in the system without requiring extra input, Plan a course of treatment in advance so that treatments are readily delivered when the patient arrives; Write RT prescriptions that detail treatment techniques, fractions, and dose; Define treatment fields; Link setup fields and notes to treatment fields; Setup notes should include photos that show how to set up the patient; Track dose to specific sites; Define site breakpoints with instructions that appear when the breakpoint will be exceeded; Store treatment plan information to avoid redundant and time-consuming data entry.</p>

	Two OIS workstation are required, therefore Bidder include the price of the same in the High Energy Linear Accelerator, no need to provide the unit price for TPS and OIS workstation offered separately.
During the warranty period, all the software updates and upgradation should be provided without asking for free of cost.	During the warranty period, all the software/hardware updates only should be provided without asking for free of cost

All other terms and conditions of the tender enquiry document shall remain unchanged. Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the newspapers.

**Sr. CGM-I, HSCC (I) Ltd**  
**For & on behalf of Director CNCI, Kolkata**