

AMENDMENT No. X – Dated 15/3/2018

Sub.: Procurement of Medical Equipment for Kalpana Chawala Government Medical College, Karnal.

Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2017-06/16 uploaded on 14.12.17

Bidders are requested to note the following technical Amendment :

Item No. 15 Semi Automated Tissue Processor for Tissue Processing

- 1 Tissue processor should be Carousal / Linear Type with min 12 station and at least 03 Paraffin baths. Instrument should have function of Vacuum/ pressure, Temp and Fume control system at all stations.
- 3 Capacity of equipment should be min 120 Tissue cassette independently for similar or different protocol.
- 4 There should be ergonomic control panel with LCD and inbuilt diagnostic software.
- 5 There should be programmable delayed start time and end time.
- 6 Cassette basket should spin counter clockwise and clockwise within the container or through Magnetic stirrer.
- 7 provided with UPS/ Inverter with 12 hrs back up.

Bid sale, submission and opening date of item No. 3,4,5,12,13,15 has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	22.03.18, 3.00 P.M.
ii.	Closing date & time for receipt of tender	22.03.18, 3.30 P.M.
iii.	Time and date of Opening of Tenders	22.03.18, 4.00 P.M.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

S/d
CGM, HSCC India Limited
For and on behalf of DGMR, Panchkula