

All Bidders**Amendment-X**

Subject: Execution including Supply, Installation, Testing & Commissioning of Modular & Minor Operation Theatres for Advanced Neuro Science Centre at PGIMER, Chandigarh.

Tender No. HSCC/SES/MOT/PGI/Neuro Science/2022 dated:24.06.2022

This has reference to above tender.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Reply
1.	<p>Page 6-7 of 8 Page 9 of 16</p> <p>14. OT LIGHT - Illumination Level: minimum 160,000 Lux each</p> <p>17. SURGICAL OT LIGHT WITH HD CAMERA d) Illumination Level: 160000Lux at both domes</p> <p>We request the illumination level mentioned of 160,000 lux should have variation of -5% tolerance as most of the manufacturers worldwide commit 160,000 lux but there is always a variation of around -5%.</p> <p>So, we request the illumination level of 160,000 lux should have -5% tolerance.</p>	Tender terms & conditions prevail.
2.	<p>Page 15-16 of 16</p> <p>The lowest responsive bidder shall submit the Manufacturer's Authorization letter as per format enclosed at Annexure-M along with quality certificates as per the standards mentioned in the tender documents i.e., BIS/US FDA/European CE Certified with 4 digit notified body number for the following equipment/items:</p> <p>A. OT Light with HD Camera B. Pendants for Anesthetist and Surgeon C. Laminar Air Flow System D. Wall & Ceiling system (SMS)</p>	Tender terms & conditions prevail.

	<p>E. Touch Screen Control Panel F. Isolation Transformer</p> <p>Sir, kindly refer to Amendment -VII:</p> <p>C. Laminar Air Flow System F. Isolation Transformer</p> <p>These two items Comes under Class I Category product, wherein BIS/US FDA/European CE Certified with 4 digit notified body number is Not Applicable. These two items should be CE Certified and does not require any Manufacturer Authorisation Letter.</p> <p>We request you to kindly amend.</p>	
3.	<p>Common</p> <p>Sir, it is requested to kindly extend the Tender Submission date from 19.10.2022 to 28.10.2022 considering the festive season.</p>	<p>Bid submission date is extended till 03.11.2022.</p>
4.	<p>BIS/US FDA/European CE Certified with 4 digit notified body number for the following equipment/items:</p> <p>a) OT Light with HD Camera b) Pendants for Anesthetists and Surgeon c) Laminar Air Flow System d) Wall & Ceiling system (SMS) e) Touch Screen Control Panel Panel f) Isolation Transformer</p> <p>Above all items are Class – I Device.</p> <p>These are Class – I device & UL/European CE Certified with 4 digit notified body number is not applicable.</p> <p>These clauses were not mention in tender document earlier & these points have been deleted in all prestigious tender / institutions. This will limit the tender for two companies' participation & again promoting cartelization.</p> <p>We request you kindly delete these points & let it be like to earlier.</p> <p>So that, this will enable wider participation in this tender, ensuring a healthy & fare competition.</p>	<p>Tender terms & conditions prevail.</p>

5.	<p>US FDA/European CE Certified with 4 digit notified body number for the following items:</p> <p>a)Wall & Ceiling Panel b)Laminar Airflow c)Touch Screen Control Panel d) Isolation Transfer</p> <p>Since these products are available indigenously with same specifications, hence as per Government notification for preference for 'Make in India 'products, while you have asking for all the imported items(US FDA/UL/European CE with 4 digit notified body number) as BIS is not applicable for these products.</p> <p>You are unnecessarily escalating the price & restricting the tender for two companies,We request you to kindly consider Indian products & They are quality products installed in all PMSSY Hospital Pan India.</p>	Tender terms & conditions prevail.
6.	<p>Bid Date Extension</p> <p>Due to festive season & bank holidays, We request you to kindly extend Bid participation date at least 10-15 Working days.</p>	Bid submission date is extended till 03.11.2022.

All other terms & conditions remain unchanged.

Chief General Manager (PG-I),
HSCC (India) Ltd.