

**HSCC (I) Ltd.**

**Amendment – XI      dated 04.12.2017**

**Tender Enquiry No. : HSCC/KCGMC/Medical Equipment/2015-06/15 dated: 15/09/2017**

Bid sale, submission and opening date for all items has been extended as per details given (Except the items No.01,03,10,11,12,15) in Table

<b>Sl. No.</b>	<b>Description</b>	<b>Previous</b>	<b>Revised</b>
1	Sale date of the tender	04.12.17, 3.00	11.12.17, 3.00
2	Closing date & time for receipt of	04.12.17, 3.30	11.12.17, 3.30
3	Time and date of Opening of	04.12.17, 4.00	11.12.17, 4.00

Bidder are also requested Technical Amendment in the Item 2 in the specification of X Ray Machine

<b>Sl. No.</b>	<b>Specification asked in the tender</b>	<b>Amended as</b>
	<b>X Ray Generator</b> e. Ma range : 125 to 800 Ma g. Ma Range : 2 to 800Ma	<b>X Ray Generator</b> e. Ma range : 125 to 500 Ma g. Ma Range : 2 to 500Ma

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

s/d

**CGM, HSCC India Limited**  
**For and on behalf of DGMR, Panchkula**

