

All Bidders**Amendment –XI**

Project: Supply, Installation, Testing & Commissioning of CSSD for Super Specialty Block at Indira Gandhi Medical College, Shimla, H.P.

IFB No. HSCC/SES/CSSD/SSB/Shimla/2019 dated 13.12.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender Document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Amended as
1.	<p>Horizontal Double Door Autoclave approx. 850-1000 litres or more litres (15 STU or More) With Accessories:</p> <p>Please make changes:</p> <p>The Tender specifications were for 12-15 STU and now with amendment No. VII, this has been made more specific with asking for just 15 STU's or more, we would request you to kindly amend capacity as 840 litres instead of 850 litres, with 12-15 STUs, which is as per original tender requirement.</p> <p><u>Amendment-X dated: 06.02.2020</u></p> <p>840 -1000 litres with 12-15 STUs with all Accessories.</p>	Deleted.

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.
For and behalf of Principal, IGMC, Shimla