

**Amendment no XII Dated 29.08.2018**  
**HSCC/KCGMC/Blood Bank Equipment/H dt. 01.05.2018**

**Procurement of Medical Equipment under KCGMC, Karnal**

All bidders are requested to note the following:

**Item no. 10, Refrigerated Centrifuge**

<b>Tender Specifications</b>	<b>May Please Read as</b>
<b>Terms &amp; Conditions</b> 2. The equipment should meet the standards of IEC/EN 61010 safety requirements for electrical equipment, for <b>measurement</b> , control and laboratory use- <b>particular requirement for laboratory centrifuges</b> . Certificate regarding the same for equipment and quoted model should be submitted from a notified approved <b>by IEC for the same</b> ; failure to which shall be disqualification.	2. The equipment should meet the standards of <b>IEC 60601/61010 as applicable for</b> safety requirements for electrical equipment, for measurement, control and laboratory use-particular requirement for laboratory centrifuges. The Certificate regarding the same for equipment and quoted model should be submitted from a notified <b>agency(national/international); the agency should be approved by IEC or Govt. of India to issue the certificate valid authorization/approval certificate of agency by the IEC or Govt. of India should be attached along with</b> ; failure to which shall be disqualification.

All other terms and conditions of the tender enquiry document shall remain unchanged. Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the newspapers.

s/d  
**Sr. CGM-I, HSCC (I) Ltd.**  
**For & on behalf of DGMER, Panchkula**