

# HSCC (India) Ltd

## AMENDMENT – XII

Dated: - 26.11.2018

**Subject:** Amendment to the tender Enquiry Document.

**Ref: Tender Enquiry No.:** HSCC/PUR/AIIMS/RAEBARELI/OPD/2018/1 dated 13.08.18

**Bid sale, submission and opening date for item no. 1 has been extended as per details given in Table -1:**

Table -1

| Sl. No. | Description   | Revised Schedule      |
|---------|---|-----------------------|
| i.      | Sale date of the tender                                 | 03.12.2018, 3.00 P.M  |
| ii.     | Closing date & time for receipt of tender               | 03.12.2018, 3.30 P.M. |
| iii.    | Time and date of Opening of Techno – Commercial Tenders | 03.12.2018, 4.00 P.M. |

**Bidders are requested to note the following technical amendment in the item no.1 :**

| Sl. No. | Tender Specification   | Amended As   |
|---------|--|--|
|         | <b>Main Unit</b><br>It should have 3 or more Storage Drawers with one of the drawers have UV sterilization. The unit should have a mirror warmer and dispenser for cotton and paper. | <b>Main Unit</b><br>It should have 3 or more Storage Drawers. The unit should have a mirror warmer and dispenser for cotton and paper. |
|         | <b>S. No.1</b><br>Layringoscope 70deg, Enlarged view, Diameter 11 mm, Working Length 185-190, Autoclavable -01 Nos.  | <b>S. No.1</b><br>Layringoscope 70deg, Enlarged view, <b>Diameter 10-11 mm</b> , Working Length 185-190, Autoclavable -01 Nos.         |

## HSCC (India) Ltd

---

|  |   |   |
|--|---|---|
|  | <p><b>S. No.4 Warm water Ear Syringing system</b></p> <ul style="list-style-type: none"><li>• In built ear rinsing system with 37 degree C temperature with chrome plated handle with autoclavable tips with quick connection.</li><li>• Separate Stainless Steel container to prevent built up</li><li>• The system should have water supply tank (Capacity 5Ltrs or more) for distilled water safety positioned for easy refilling.</li><li>• Ear rinsing funnel</li><li>• Two sets of rinsing cannula (3 sets)</li></ul> | <p><b>S. No.4 Warm water Ear Syringing system</b></p> <ul style="list-style-type: none"><li>• In built chrome plated autoclavable handle</li><li>• Separate Stainless steel container to prevent mineral built up</li><li>• Should include sinus drainage tube, several front cannulas(luer) and water supply hose including water filter with cartridge</li><li>• Electronic Temp control of water temp</li><li>• Ear rinsing funnel and automatic liquid discharge/ drainage system</li></ul> |
|--|---|---|

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

Director (PMSSY)  
Ministry of Health and Family Welfare