

Amendment no XIII Dated 19.04.2018
HSCC/PUR/CNCI/Kolkata/Medical Equipment/02 dt. 02.11.2017

Procurement of Medical Equipment CNCI 2nd Campus

All bidders are requested to note the following:

Item No	NAME OF THE EQUIPMENT	Last date & time sale/downloading of Tender document	Last closing/submission date & time for receipt of tender	Date of opening of Techno – Commercial bids.	Earlier date of opening	Amendment Status for opening date
4	PET – CT (Positron Emission Tomography CT Scan)	26.04.2018, up to 13.00 hrs IST	26.04.2018, 14.00 hrs IST	26.04.2018, 14.30 hrs IST	16.04.2018	Technical Specification updating

Item No.4, PET CT

Tender Specifications Item No.4, PET CT	Bidder representation	Amended as
Water proofing of the whole roof by Brick Bat COBA. Water proofing treatment of the mother slab should also be done with tapecreate (CICO) chemical and 12 mm thick cement plaster protecting layer	Building is under construction and HSCC get it done through civil vendor (as per discussed at site)	PET CT was placed on first floor hence Water proofing of the whole roof by Brick Bat COBA is not applicable.
Air-Conditioning specification (highlighted in attached file)	Specification are not feasible for site (VRV system of capacity 84 HP (28HP x 3 nos) plus additional ductable AC 17TR (8.75 X 2TR) asked in the tender which needs to be corrected for area 793 sqft layout attached along with this letter.	Proper Air conditioning system with suitable capacity should be provided for the designated area

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

Chief General Manager, HSCC (I) Ltd
For & on behalf of Director CNCI, Kolkata