

HSCC I Ltd

Amendment –XIV

Dated 15.02.2018

Tender Enquiry No: HSCC/SJH/Medical Equipment/35 Dated 18.10.2017

Tender Enquiry No: HSCC/SJH/Medical Equipment/35 Dated 18.10.2017

The bid submission date may be extended as per table –I for Item No. 2, 5, 8

Table: I

Sr. No.	Description	Detail of Items	Previous Date	Revised Date
i.	Sale Date of the tender	Item No. 2, Bi-Spectra Index Monitor (Non EFC)	16.02.2018, 2.30 PM	23.02.2018, 2.30 PM
ii.	Closing Date & Time for receipt of Bids	Item No. 5, Hemodialysis Machine with Portable RO(Non EFC)	16.02.2018, 2.30 PM	23.02.2018, 2.30 PM
iii.	Time and date of Opening of Tender	Item No. 8, Angiography System for Interventional Hybrid Procedure, (Non EFC)	16.02.2018 , 3.00 PM	23.02.2018 , 3.00 PM

For Item No. 8, Angiography System for Interventional Hybrid Procedure, the technical amendment & Approved layout plan have been received. It is therefore extended from 16.02.2018 to 23.02.2018. for Item No.8 Angiography System for Interventional Hybrid Procedure and site layout plan are as under:

Item No. 8

SPECIFICATION FOR ANGIOGRAPHY SYSTEM FOR INTERVENTIONAL AND HYBRID PROCEDURES

INCLUDING TAVI

State of the art, single floor/ceiling mounted C-arm/G-arm system with flat detector technology digital imaging system for hybrid, interventional and TAVI procedure in Hybrid OR environment.

A) C-ARM/G-ARM MULTI-DIRECTIONAL FLOOR/CEILING MOUNTED:-

1. All movements should be motorized with C-Arm angulations of minimum RAO/LAO+100 deg. /- 100 deg. CRAN/CAUD + - 45 deg. At head end position. With 20 deg. /sec. or more speed for LAO/RAO and 15 deg./sec or more speed for CRAN/CAUD.
2. The system should have user defined 50 programmed position of the C-Arm and table.
3. Motorized parking of C-Arm away **or left or right** from table for operations on table when C-Arm not in use. C Arm mounting/base should not come in the area of the operation table and operating area.
4. The C-arm should have auto collision protection with patient and the table.
5. C-arm must not interfere with laminar airflow **during surgery** and accommodate planer of min size of 2400x2400 cm on ceiling and OT lights on table in working position or while moving the C arm to different positions.

B) TABLE:-

- 1 Floor mounted table with flat table with radiolucent table top and floating functions with rotation from the base.
- 2 Table should be integrated with the system and should have motorized vertical movement.
- 3 The table should have Integrated Emergency Stop; all motorized movements (including table), are stopped when the System emergency stop button is pressed. The table should have Integrated Collision detection, all motorized movements (including table), are slowed down or stopped when detects the patient Workflow.
- 4 The table should have motorized axis tilt (head-down and head-up as well lateral tilt) with at least 15 deg.
- 5 The system should have provision for collision protection.
- 6 Table should support wide range of accessories like arm rest, surgical screen, instrument tray, hand grips, and one IV pole.
- 7 Gantry controls, table system controls, Collimation controls to be mounted on trolley away from table for better flexibility.
- 8 Table should support patient weight up to 200Kgs or more & should take full patient weight during CPR in extended table position.
- 9 The table should have head to toe coverage without need for repositioning of the patient . Please specify the table dimensions.
- 10 System should have well designed & light weight footswitch for releasing fluoroscopy, acquisition and table brakes.

C) X-RAY GENERATOR:

- 1 100 KW or more compatible with high resolution imaging
- 2 High frequency X-ray generator with automatic regulation of radiation dose rate for all fluoro DSA and acquisition imaging.
- 3 The minimum power rating should be 100 KW OR more at 100 KV compatible with high resolution imaging.
- 4 Fluoroscopy exposure and mA should be automatically controlled.
- 5 System should have pulsed fluoroscopy system.

D) X-RAY TUBE:

- 1 X-Ray tube should be with fine focal spots (2/3 focal spots for different application) with high cooling rate to ensure continuous operation, capable of pulsed fluoroscopy. The large Focus power output should be 65 kw to 80 kW or more. The pulse Fluoroscopy should be offered with pulse rate of 3.75 Frame/sec to 30 frames/sec.
- 2 The X-ray tube should have Anode heat storage capacity of at least 2.4 MHU or more to run continuously for 6-8 hours without shutting off **with optional secondary grid switch technology.**

E) RADIATION PROTECTION:

- 1 The system should have integrated computer controlled X-ray Beam filtering with at least 3/4 programmable copper filters of various sizes from 0.2 mm to 0.9 mm.
- 2 The system should have positioning of collimator blades without radiation.
- 3 The system should have monitoring and display of X-ray dose during the patient examination. It should be possible to create a DICOM based dose report of the patient.
- 4 System should meet all National & International safety standards, US FDA , European CE & comply with BARC & AERB guidelines.

- 5 System should have latest low dose radiation safety features of operator and patient (care & clear /clarity IQ or equvalant)

F) DIGITAL IMAGING SYSTEM:

- 1 A flat detector with a diagonal size of least 46 cm. **with 16 bit digitalization depth.**
- 2 Digital system with acquisition and processing in 1024 x 1024 matrix at 25/30 fps with 8/10/12 bit digitization.
- 3 Image storage capacity of at least 1,00,000 images in 1024 x 1024 matrix at 10/12 bits on the main system disk and upgradeable further.
- 4 System should have capability of ECG display on the live image monitor during the cina acquisition and perferably archive along with angio images on CD, during the acquisition.
- 5 System should have on-line & off-line validated **cardiac and** vascular analysis. The software should have Auto calibration facility for Stenosis measurement with geometrical and densitometry calculations. The analysis should be possible from table side in the examination room and from the control room.
- 6 The full system should have table side control operation with touch screen / knobs for complete acquisition and post processing capabilities.
- 7 The system should have on-line DSA capabilities with acquisition frame rate of 1 frame/sec to 6 frames /sec or more.
- 8 Rotational angiography facility with and without subtraction should be available. It should be possible to generate 3D and CT like images from rotational angiography data.
- 9 The system should have facility for storage of fluoro loop scene of at least 10 seconds.
- 10 System should be provided with all advanced 2D and 3D road mapping.
- 11 The latest complete software and hardware for visualizing stent with extra high- resolution from table side control & preferably stent enchancement with relation to lumen (fade in and fade out).
- 12 Separate work station should be provided with at least 3 GB main memory for 3D reconstruction to free the main system for continuation of procedure immediately after displayed both in the control room as well as the examination room. It should be possible to recall & view images acquired in the rotational mode (subtracted &un subtracted) alongside the 3D images. Review of all images should be fast interactive & user friendly in both the rooms. Table side controls for 3D reconstruction and C-Arm positioning with respect to 3D image & selection of 3D image with respect to C-Arm positioning should be provided. The 3D reconstruction should be in true 512 matrix. All options related to 3D (MIP, SSD, VRT, slicing, measurement tool, volume try etc.) should be available. Software to fuse CT, angio & MR images of the same patient at the independent workstation is desirable.
The 3D workstation should be capable of the following functions
 - i) 3D reconstruction using SSD, VRT, MIP in true 512 matrix with color display in the control & examination room.
 - ii) All post-processing functions for images including fusion of CT, MR and angio images with 3D display & 3D volume measurement.
 - iii) Immediate background transfer of all images as soon as they are acquired, to CD/DVD recording station.
 - iv) It should be possible to view dual density objects in one view to differentiate blood vessels from coils.
 - v) It should be possible to have 3D of aortic root from the rotational angio data/ pre acquired CT data and auto marking of anatomical landmarks like coronary ostium. It should be possible to have overlay of live fluoro on this 3D image or valve marking lines with landmarks for guidance in TAVI procedures. C Arm should position itself automatically depending upon the 3D image.

G) MONITORS/DISPLAY:

- 1 The monitor display system in examination room should be ceiling suspended and it should be possible to position it on the left or right side of patient table. The monitor should be single high resolution monitor of at least 55" and 8 megapixel resolution with luminance of 700 CD / M. sq and 1000:1 contrast or more to display live,reference,3D image ,hemodynamic data . Two additional monitors of at least 19" to be provided. One for live image and another for waveform display.
- 2 Console room should have monitors to display patient demographics, fluoroscopy images and 3D images. All monitors at least 19 " or more
- 3 **Hemodynamic specification:**
 - a. dedicated hemodynamic system for cardiac catheterization for adult and peadiatric patient
 - b. Hemodynamic system should have four invasive pressures & ECG.
 - c. Should have all Hemodynamic analysis tools to calculate cardiac output, shunt area and Gradient measurements
 - d. Should be able to store and print one week angio-studies.
 - e. 12 lead ECG and 4 lead Invasive pressure transducer and domes should be supplied with the system.
 - f. Should be connected with cath lab network printer.
 - g. 20 personal dosimeters should be provided which will give real time reading of the radiation dose.
 - h. There should be a provision for display of time taken for entervential procedures in the operating room the dosimeter should have wired connectivity to the dose monitoring system for according the doseages and archiving

H) CONTROL CONSOLE:

- 1) All system movements of C-Arm, table, image display, image review, image post processing and quantification shall be controlled both by the operator at the table in the exam room and in console room.
- 2) The system should have facility for edge enhancement, Positive/negative image display, windowing, contrast/brightness, electronic shuttering, image/pixel shifting, vertical and horizontal image reversal, zoom functions. System should have facility to allow overlying of fluoroscopy and reference image with fade-in and fade-out of these images possible.
- 3) The system should have fast and direct access to all series, single images, in both examination and control room.
- 4) System should have angle/distance measurement, Image labeling and patient positioning facilities.
- 5) **May have optional Fade in Fade out facility.**

I) DIGITAL ARCHIVING:-

- 1 System should be DICOM 3.0 ready.
- 2 Image transfer from digital system in background mode without affecting the system operation.
- 3 USB interface to copy images to memory disk/external hard disk.

J) IVUS

1. Latest Generation Intra Vascular Ultrasound System 2.7 version.
2. Monitor-SVGA LCD Monitor-Minimum 19".
3. Colour map (iMap) overlay to visualize tissue type (Fibrotic, Lipidic, Necrotic, and Calcified) as well as an improved workflow design.
4. Area & length measurements graphics in the cross-sectional and long view images.
5. ECG and Audio signals capturing enabled.
6. Ease to eyes by colouring IVUS images.
7. Sterile field control option.
8. Compatibility with Coronary, **Peripheral** and optional intra cardiac echo catheters (**ICE**).
9. Automated lumen and vessel measurement to aid in diagnosis and planning.
10. Dynamic review feature (blood flow, plaque morphology, dissections and stent apposition)
11. Long view and cross sectional imaging.
12. Coronary IVUS catheter with single Rotating/Mechanical transducer driven by a flexible drive cable with minimum 20 MHz frequency.
13. High storage capacity with removable hard disk and minimum of 25 cases storage.
14. Archiving options: CD ROM, 16X DVD, Removable Hard Disk and Network.
15. Maximum number of area measurements per cross section image for better imaging, Minimum-3d
16. Maximum number of distance measurements per cross section images for better imaging, Minimum-9
17. Minimum 30 GB hard disk with an option of removable storage with minimum storage of 20 cases or more.
18. DICOM storage & image formatting.
19. Digital Frame Grabber.
20. Data Entry: Touch Screen, Mouse
21. Multiple image screen format.
22. Automatic & manual pull back options.
23. Automated vessel and lumen measurements/detection with manual correction option.
24. Printer to print IVUS images.
25. Approval-International standards agency US FDA or Equivalent.
26. Clinical support for training of staff.
27. Technical back up for maintenance of machine

K) FFR (Fractional Flow Reserve) Measurement System for Physiological Lesion Assessment

1. Should be able to assess Arterial Pressure through Cath Lab transducer system wirelessly/ with wire.
2. Should be able to assess distal Pressure through device, both wired /wirelessly.
3. Should have the option to be wall mounted, table mounted or desktop as per the need.
4. Should be able to transmit and display recordings on the existing cath lab monitors
5. Display both real time pressure and mean pressure values.
6. Screen window displays real time FFR in both numerical and graphical form.
7. Should have optional upgraded software to calculate CFR (Coronary Flow Reserve) this is real time. Both FFR and CFR simultaneously. IMR to be calculated using desktop software or should have optional upgraded software for instantaneous flow reserve (IFR) & IFR scout calculating software.
8. Should be capable of reading temperature (thermodilution)
9. Temperature readings in the range of 15-42 C
10. Should be able to do Pressure Reference equalization

11. Should measure pressure in a range of -30 to 300 mm Hg
12. Touch screen capabilities to be operator friendly, guide steps to follow for procedure.
13. Displays Calibration steps.
14. Should have a frequency response of 0-25Hz
15. Should give Graphical presentation of pressure waves.
16. Allows different beat settings in accordance with the cath lab system.
17. Should be compatible with different hemodynamic systems
18. Should have memory to save and record the data.
19. Should also enable cath lab to read AO pressure for monitoring.
20. Should communicate with PC.
21. Should be compatible with printer, DICOM and USB interface for transferring FFR recordings

L) **Optical Coherence Tomography (OCT) System with real time online 3D imaging features**

1. The system should have an imaging engine that is based on the fiber optic technology.
2. The system should have wireless FFR measurement capabilities.
3. It should utilize catheter that emit near infra red light to produce high resolution real time images.
4. Should have two monitors plus remote video output for multiple
 - a. sightlines.
5. The system should have an integrated drive-motor and Optical Controller (DOC).
6. Should have an isolation transformer.
7. Should have a computer, a keyboard, and a mouse.
8. CPU with high end DAS card for faster 3-D data acquisition speed
9. 22*CD/DVD RW dual player DVD RAM drive for faster image management.
10. DICOM compatibility

The system should allow the user to :

1. Acquire, save and subsequently retrieve images for review. **Real-time 3D image**
 - a. **Re-construction of lumen and vessel**
2. Immediate and accurate lumen boundary detection and Lumen Profile Display
3. Stent planning workflow with automated minimum lumen area and percent stenosis measurements
4. Automatic lumen detection on every frame
5. Profile of mean diameter or lumen area across pullback
6. Automatic marking of MLA frame
7. User-defined proximal and distal reference frames
8. Automated display of reference frame area and diameters, distance between references, %AS and %DS
9. Automated measurements mode for calculations for stent sizing
10. Seamless integration of FFR and OCT with guided workflows for exceptional ease-of-use
11. Should allow user for easy orientation on Angiography
12. Allow to acquire and review images in L-Mode (lateral view).
13. Overlay color maps to optimize contrast resolution.
14. Enlarge a defined area of interest (zoom).
15. Make measurement and calculations of % Diameter stenosis
16. Add text annotations.
17. Play back and edit images with a full range of playback and editing capabilities.
18. Export still images and movies in raw OCT format or in standard AVI, TIFF, JPEG,BMP, or DICOM formats.
19. Import OCT format images and review and edit them with full OCT

- a. review and edit capability.
20. Perform basic file management functions.

The imaging Parameters of the system should be:

1. Maximum frame rate: Up to 180 fps
2. Longer pullback of up to 75 mm and up to 540 frames
3. Faster pullback speed up to 36 mm/sec
4. Allows user to do high resolution imaging for online real time 3-D re-construction
5. # of lines per frame: 500
6. Scan diameter:10 mm
7. Axial Resolution: 15 microns

M) Training

Training to be provided to six doctors/staff of the department for operation & maintenance of the equipment for two weeks at an International centre of repute abroad with similar equipment installed.

N) UPS:-

Suitable online UPS of at least 120 KVA capacities with 30 min. battery backup for complete Cath Lab including cine and fluoroscopy. Emergency lighting should also be on UPS

O) ESSENTIAL ACCESSORIES TO BE SUPPLIED:

- 1 Lead glass 150 x 120 cm.(as per international radiation protection standard)
- 2 Good quality, wrap around light weight Lead Aprons with hangers-20 nos. (as per FDA standard)
- 3 Thyroid Guard-20 nos. (as per international radiation protection system)
- 4 Ceiling suspended radiation protection-1 no. (as per international radiation protection system)
- 5 Table mounted radiation protection -1 no. (as per international radiation protection system)
- 6 Integrated two way communication system between control room and examination room.
- 7 One laser Network Printer of high resolution (at least 1200 dots per inch) with minimum 128 MB memory and 1200 dpi
- 8 High –pressure injection system, having features of programmed flow rate, volume. Vendor should supply 200 syringes

Site Modification:

1. Area under the scope of the site modification is :-
 - o Examination Room
 - o Console Room
 - o UPS cum Cabinet room
2. Examination room should be in line with AERB requirements
3. Lead Glass window and lead door to be provided on console and examination room common wall. Second lead lined door to be provided in the examination room for patient entrance
4. PVC Flooring, trenches and cable trays to be provided.

5. Air-conditioning of 2 x 2 TR(Split AC) to be provided in the UPS cum cabinet room
6. Total load of the equipment along with the accessories to be mentioned by the vendor
7. Hospital authorities to provide electrical cable upto electrical panel in the UPS room
8. LT panel should be Supplied & installed by the vendor

9. All general lighting, electrical fittings and fixtures to be provided by hospital authorities
10. Adequate fire detection system to be provided by the vendor, inside the area under the scope mentioned above
11. Partition to be provided between UPS and Batteries in the UPS room
12. All civil works related to successful installation of the machine as per the layout attached should be carried out by the vendor
13. Wall finishes in console and UPS room - upto fall ceiling skirting along with POP and plastic emulsion paint upto the false ceiling.


Furniture:

1. Workstation table – 1 no.
2. Computer chairs on castors – 2 nos.
3. Cupboard with laminated door shutters (1200 mm x 600 mm x 2100 mm) – 2 nos.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website /CPP Website for corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regards.

Medical Superintendent
Safdarjung Hospital New Delhi



FIRST FLOOR PLAN SHAFT SCHEDULE - ZONE-2

TAG CODE	DESCRIPTION	LENGTH	WIDTH
ELEC SH-07	ELECTRICAL SHAFT	300	1000
ELEC SH-08	ELECTRICAL SHAFT	500	550
ELEC SH-09	ELECTRICAL SHAFT	500	550
ELV SH-02	ELV SHAFT	750	300
HVAC SH-15	HVAC SHAFT	350	1200
HVAC SH-20	HVAC SHAFT	6750	1050
HVAC SH-22	HVAC SHAFT	3000	1600
HVAC SH-56	HVAC SHAFT	600	2200
HVAC SH-59	HVAC SHAFT	600	2000
HVAC SH-74	HVAC SHAFT	500	1300
HVAC SH-100	HVAC SHAFT	1500	750
HVAC SH-105	HVAC SHAFT	650	655
MGPS SH-01	MGPS SHAFT	600	1000
PHE SH-22	PHE SHAFT	2400	750
PHE SH-23	PHE SHAFT	2000	750
PHE SH-24	PHE SHAFT	1155	700
PHE SH-26	PHE SHAFT	300	1600
PHE SH-27	PHE SHAFT	1200	1500
PHE SH-28	PHE SHAFT	600	850
PHE SH-29	PHE SHAFT	2435	870
PHE SH-30	PHE SHAFT	600	2100
PHE SH-32	PHE SHAFT	450	600
PHE SH-33a	PHE SHAFT	900	750
PHE SH-64	PHE SHAFT	1315	485

WALL LEGEND:

- 230MM OR 155MM THK. 2hrs FIRE RATING BRICK WALL UPTO SOFFIT OF BEAM / SLAB
- 230MM OR 155MM THK. BRICK WALL UNHATCHED SHALL BE UPTO SOFFIT OF BEAM / SLAB
- 230MM OR 100MM THK. AAC BLOCK WALL UPTO SOFFIT OF BEAM / SLAB
- 230MM OR 155MM THK. BRICK WALL UPTO 2100MM HEIGHT
- 230MM OR 155MM THK. BRICK WALL UPTO 2800MM HEIGHT
- 200MM OR 100MM THK. AAC BLOCK WALL UPTO 3100MM HEIGHT
- 230MM OR 155MM THK. BRICK WALL UPTO 1500MM HEIGHT (EDGE WALL)
- RC WALLS / COLUMNS

SUNKEN LEGEND:

- STRUCTURAL SLAB SUNKEN BY 50MM FROM FFL

FIRST FLOOR PLAN DOOR SCHEDULE - ZONE-2


TAG	WIDTH	HEIGHT	DESCRIPTION	COUNT
AD09	1500	2100	POWDER COATED ALUMINIUM FRAMED PARTIALLY PANELED DOOR OF TWO EQUAL SHUTTERS	2
AD10	1200	2100	POWDER COATED ALUMINIUM FRAMED FULLY PANELED SINGLE SHUTTER DOOR	1
AD11	1600	2100	POWDER COATED ALUMINIUM FRAMED PARTIALLY GLAZED AND PARTIALLY PANELED SINGLE SHUTTER DOOR	7
AD13a	750	2100	POWDER COATED ALUMINIUM FRAMED FULLY PANELED SINGLE SHUTTER DOOR	1
AD15	1800	2100	POWDER COATED ALUMINIUM FRAMED PARTIALLY GLAZED PARTIALLY PANELED SINGLE SHUTTER DOOR	2
AD17	600	1800	POWDER COATED ALUMINIUM FRAMED FULLY PANELED SINGLE SHUTTER DOOR	3
AW002	1350	2100	POWDER COATED ALUMINIUM FRAMED WOODEN FLUSH DOOR OF TWO EQUAL SHUTTERS	2
AW005	1200	2100	POWDER COATED ALUMINIUM FRAMED WOODEN FLUSH DOOR OF TWO EQUAL SHUTTERS WITH VISION PANEL-200X300MM	3
AWD11	1050	2100	POWDER COATED ALUMINIUM FRAMED WOODEN FLUSH DOOR OF SINGLE SHUTTER WITH VISION PANEL-200X300MM	3
AWD11a	1050	2100	POWDER COATED ALUMINIUM FRAMED WOODEN FLUSH DOOR OF SINGLE SHUTTER WITH VISION PANEL-200X300MM	6

FIRST FLOOR PLAN WINDOW SCHEDULE - ZONE-2

TAG	WIDTH	HEIGHT	DESCRIPTION	COUNT
AW2a	1500	1500	POWDER COATED ALUMINIUM FRAMED OPENABLE DOUBLE GLAZED WINDOW OF TWO EQUAL SHUTTERS WITH ANODIZED ALUMINIUM SUB FRAME & ALUMINIUM LOUVERS ABOVE	3
AW04a	1000	1500	POWDER COATED ALUMINIUM FRAMED FIXED DOUBLE GLAZED WINDOW OF SINGLE SHUTTER WITH ANODIZED ALUMINIUM SUB FRAME	1
AW09	1600	1500	POWDER COATED ALUMINIUM FRAMED OPENABLE DOUBLE GLAZED WINDOW OF THREE EQUAL SHUTTERS WITH ANODIZED ALUMINIUM SUB FRAME	1
AW08a	1800	1500	POWDER COATED ALUMINIUM FRAMED FIXED DOUBLE GLAZED WINDOW OF THREE EQUAL SHUTTERS WITH ANODIZED ALUMINIUM SUB FRAME	3
AW11	1500	1500	POWDER COATED ALUMINIUM FRAMED OPENABLE DOUBLE GLAZED WINDOW OF TWO EQUAL SHUTTERS WITH ANODIZED ALUMINIUM SUB FRAME	2
AW11a	1500	1500	POWDER COATED ALUMINIUM FRAMED FIXED DOUBLE GLAZED WINDOW OF TWO EQUAL SHUTTERS WITH ANODIZED ALUMINIUM SUB FRAME	1
UW1	1800	1500	UPVC FIXED WINDOWS OF THREE EQUAL SHUTTERS WITH DOUBLE GLAZING & ANODIZED ALUMINIUM SUB FRAME	3
UW1a	1800	1500	UPVC OPENABLE WINDOWS OF THREE EQUAL SHUTTERS WITH DOUBLE GLAZING & ANODIZED ALUMINIUM SUB FRAME	1
UW2	1500	1500	UPVC FIXED WINDOWS OF TWO EQUAL SHUTTERS WITH DOUBLE GLAZING & ANODIZED ALUMINIUM SUB FRAME	3
UW2a	1500	1500	UPVC OPENABLE WINDOWS OF TWO EQUAL SHUTTERS WITH DOUBLE GLAZING & ANODIZED ALUMINIUM SUB FRAME	1
UW3	1200	1500	UPVC FIXED WINDOWS OF TWO EQUAL SHUTTERS WITH DOUBLE GLAZING & ANODIZED ALUMINIUM SUB FRAME	1
UW4	1200	1200	UPVC OPENABLE WINDOWS OF SINGLE SHUTTER WITH DOUBLE GLAZING & ANODIZED ALUMINIUM SUB FRAME	1

FIRST FLOOR ZONE-2 STRUCTURAL GLAZING SCHEDULE

TAG	WIDTH	HEIGHT	DESCRIPTION	COUNT
S06	4150	1500	STRUCTURAL GLAZING	1
S07	11250	1500	STRUCTURAL GLAZING	1
S08	6750	2400	STRUCTURAL GLAZING	1



- GENERAL NOTES:**
- ALL DIMENSIONS ARE IN MILLIMETERS AND LEVELS ARE IN METERS.
 - THIS DRAWING SHOULD BE READ IN CONJUNCTION WITH THE FOLLOWING DRAWINGS O13189-A-HP-01-WD-0013 & 0015.
 - FOR JOINERY, STAIRCASE, RAMP DETAILS LIFTS, TOILETS, WALL SECTIONS, FLOORING PATTERN DETAILS & FALSE CEILING DETAILS REFER SEPARATE DRAWINGS.
 - FOR STRUCTURAL PUBLIC HEALTH ENGINEERING, HVAC, ELECTRICAL, FIRE FIGHTING AND GAS PIPING DETAILS REFER APPROPRIATE SEPARATE DRAWINGS.
 - LIFT WALL OPENING SIZES AND POSITIONS TO BE AS PER LIFT MANUFACTURERS SPECIFICATIONS.
 - ALL SHAFT SIZES PROVIDED HERE ARE CLEAR DIMENSIONS AFTER DEDUCTING THE BEAM SIZES.
 - ALL INTERNAL ELEC. PHE & HVAC SHAFTS SHOULD HAVE 100MM HIGH R.C.C. THRESHOLD FROM FFL AT ALL FLOOR OPENINGS.
 - ALL EXTERNAL WALLS SHOULD BE BUILT UPTO SOFFIT OF BEAM OR CEILING.
 - ANY DISCREPANCIES NOTICED IN THIS DRAWING SHOULD BE BROUGHT TO EDRCS NOTICE IMMEDIATELY.
 - ALL TOILETS SHALL BE REFERRED TO SEPARATE TOILET DETAIL DRAWINGS.
 - REFER SEPARATE DRAWING FOR LIFT LOBBY DETAILS.
 - ALL ROOF SLAB CUTOUTS INDICATED IN THIS DRAWING ARE BEAM/CEILING. REFER STRUCTURAL DRAWINGS FOR EXACT LOCATION.
 - ALL INTERNAL WALLS SHALL BE BUILT UPTO SOFFIT OF BEAM/CEILING UNLESS OTHERWISE SPECIFIED IN THE DRAWINGS.
 - FOR FHC & HVAC SHAFT WALLS SHALL BE REFER RESPECTIVE STRUCTURAL DRAWINGS.
 - STRUCTURAL SHEAR WALLS / RETAINING WALLS / COLUMNS SIZES / SETTING OUT TO BE REFERRED TO RESPECTIVE STRUCTURAL DRAWINGS.
 - FINISHING MATRIX SHALL BE REFERRED FOR ARCHITECTURAL FINISHES IN THIS FLOOR.
 - FOR WALL OPENINGS REFER CO-ORDINATED DRAWINGS.
 - ALL RECESS DOORS SHALL HAVE LINTEL SLAB AS PER STRUCTURAL DRAWINGS COVERING BOTH INTERNAL & EXTERNAL WALLS AT 2100MM FROM FFL.
 - ALL EWC, IWC, BATH CUBICLES SHALL HAVE 800MM x 1950MM HPL DOORS AS PER VENDOR DETAILS.
 - ELEVATIONAL CLADDING MATERIALS ARE NOT INDICATED IN THE PLANS. FOR CLARITY REFER SEPARATE DRAWING FOR DETAILS.

FIRST FLOOR PLAN DOOR SCHEDULE - ZONE-2

TAG	WIDTH	HEIGHT	DESCRIPTION	COUNT
AWD12	1000	2100	POWDER COATED ALUMINIUM FRAMED FLUSH DOOR OF SINGLE SHUTTER WITH ALUMINIUM LOUVER OF SIZE 500X300MM	3
AWD13	1000	2100	POWDER COATED ALUMINIUM FRAMED WOODEN FLUSH DOOR OF SINGLE SHUTTER	1
AWD16	800	2100	POWDER COATED ALUMINIUM FRAMED FLUSH DOOR OF SINGLE SHUTTER WITH ALUMINIUM LOUVER OF SIZE 500X300MM	8
FD03	1500	2100	HMPS DOOR OF TWO EQUAL SHUTTERS 2HRS. FIRE RATING	10
FD04a	1200	2400	HMPS DOOR OF TWO EQUAL SHUTTERS 2HRS. FIRE RATING	1
FD08	1000	2100	HMPS DOOR OF SINGLE SHUTTER 2HRS. FIRE RATING	1
FD11	600	1800	HMPS DOOR OF SINGLE SHUTTER 1HR. FIRE RATING	7
FD13	1200	2100	HMPS DOOR OF TWO EQUAL SHUTTERS WITH GLASS PANEL ON TOP 2HRS. FIRE RATING	2
FD15	600	1200	HMPS DOOR OF SINGLE SHUTTER WITH 1HR. FIRE RATING	1
SL001	1500	2100	HERMETICALLY SEALED ALUMINIUM FLUSH SLIDING DOOR	6
WF001	2000	2100	TEAK WOOD FRAMED WOODEN FIRE RATED WOODEN FLUSH DOOR OF TWO EQUAL SHUTTERS WITH 200X300MM VISION PANEL - 2 HR. FIRE RATING	6
WF002	1800	2100	TEAK WOOD FRAMED WOODEN FIRE RATED WOODEN FLUSH DOOR OF TWO EQUAL SHUTTERS WITH 200X300MM VISION PANEL - 2 HR. FIRE RATING	1

REVISIONS

NO.	DESCRIPTION	DESIGNED	DRAWN	CHECKED	APPROVED
1	ISSUED FOR CONSTRUCTION	TI	PN	NNT	JA
2	ISSUED FOR CONSTRUCTION	TI	PN	NNT	JA

CHECKED BY:

SIGNATURE	DATE	CHECKED BY	SIGNATURE	DATE	CHECKED BY	SIGNATURE	DATE
[Signature]	20/10/14	HVAC	[Signature]	20/10/14	GIS	[Signature]	20/10/14
[Signature]	20/10/14	FPS	[Signature]	20/10/14	GEOTECH	[Signature]	20/10/14
[Signature]	20/10/14	PIPING	[Signature]	20/10/14		[Signature]	20/10/14
[Signature]	20/10/14		[Signature]	20/10/14		[Signature]	20/10/14

L&T Construction
Buildings & Factories

CLIENT: M/s. HSCC (INDIA) LTD. FOR M/s. SAFDARJUNG HOSPITAL
CONSULTANT: Engineering Design and Research Center (EDRC), L&T Construction - Buildings & Factories

PROJECT: Proposed Super Specialty Block cum Paid Ward, Safdarjung Hospital, New Delhi.
CONCEPT ARCHITECT: M/s. HSCC (India) Ltd.

SUPPLIER/CONTRACTOR: Larsen & Toubro Limited, Construction, Buildings & Factories

JOB NO: O13189
TITLE: SUPER SPECIALTY BLOCK CUM PAID WARD

DRAWING No.: O13189 - A - HP - 01 - WD - 0014
SECTION: ELEVATION MARKING
SCALE: 1:125
PROJECTION: FIRST FLOOR PLAN - ZONE 2
SIZE: A1
NO.: 2

RELEASED FOR: PRELIMINARY TENDER INFORMATION APPROVAL CONSTRUCTION