

Amendment No. XIV Dated 23.04.2018
HSCC/PUR/CNCI/Kolkata/Medical Equipment/03 dt. 15.11.2017

Procurement of Medical Equipment CNCI 2nd Campus

All bidders are requested to note the following:

Item No	Name of the equipment / Instrument	Last date & time sale/downloading of Tender document	Last date & time closing/submission for receipt of tender	Date of opening of Techno – Commercial bids.	Earlier date of opening	Amendment Status for opening date
1	MRI 3 Tesla	10.05.2018, up to 13.00 hrs IST	10.05.2018, 14.00 hrs IST	10.05.2018, 14.30 hrs IST	26.04.2018	Date Extension with technical Specification updating.
2	CT Scan 256 Slice					
3	Digital Mammography					
4	Digital Fluro Radiography					
5	Digital Mobile X – Ray					
6	High End Ultrasound Machine					
7	Mid End Ultrasound Machine					

Item No.3 Digital Mammography		
Tender Specifications Item No.3. Digital Mammography	Bidder representation	Amended as
Specs XII 5 conflicts with your section IX qualification criteria.: The offered unit should have a minimum of 3 installations in the country, of the same model, and this should be supported by end user certificates, purchase order copies and installation reports.	<p>A) has request: The offered unit should have a minimum of 1 installation in the country, of the same model, and this should be supported by end user certificates, purchase order copies and installation reports.</p> <p>OR</p> <p>The bidder should have supplied at least 3 units of quote model and completed installation of at least 1 unit .</p> <p>(we have supplied 3 units and completed 1 installation at TATA Memorial Hospital and other 2 units at Advanced cancer Institute and Kalapana Chawla Medical college are installed by us but handover is under process . We have 6 other mammography installations in reputed government medical colleges but of different model)</p>	The offered unit should have a minimum of 1 installation in the country, of the same model, and this should be supported by end user certificates, purchase order copies and installation reports.

Item No.4 Digital Fluro Radiography		
Tender Specifications Item No.4. Digital Fluro radiography	Bidder representation	Amended as
<p>General : High powered X-ray unit with Digital flat panel for various fluroscopy and radiography examinations for the department of radio diagnosis. The Unit should be equipped with integrated high-frequency generator, digital detector and Digital Image processing system. It should be capable of performing all plain and contrast enhanced radiology and fluoroscopy along-with angiography facility for interventional procedures. Among three major components tube/generator /detector at least two component must be manufactured by quoting vendor themselves. It should be US FDA approved. Type approval from AERB is mandatory. The vendor should have prior experience of supplying same/similar equipment in India in the reputed government or private institutions as per DGHS /MOHFW guidelines. The order copies and performance certificates from these reputed (Govt./Private institutions)should be available. The system should have the following essential features. The bidder should quote their latest model.Please mention year of launch. Supplier should have a trained service engineer in the state of supply for better uptime.</p>	<p>M/s Platinum Health care has request: Pls DELETE clause “ Among three major componentsquoting vendor themselves” as ONLY 1 COMPANY CAN COMPLY WITH THIS SPECIFICATION. We will cover all these components under 5 year warranty & 5 year CMC , it is immaterial who makes these components. We use most reputed Tixell/Thales detector (all firms use same detector) and most reputed AERB approved VARIAN OR IAE tube which are covered under 5 year full risk warranty & CMC so kindly do not restrict competition due to this clause as it suits only 1 bidder . Pls also see attached AIIMS – National Cancer Institute tender where this clause was not specified and they got 3 compliant bids.</p>	<p>Deleted Line " Among three major components tube/generator /detector at least two component must be manufactured by quoting vendor themselves".</p>
	<p>A) has request: Pls ADD “ Remote controlled compression cone is mandatory for G.I. studies. We attach latest AIIMS – National Cancer Institute Jhajhar Specs (see Table specs – page 81 , specs 1.7) for this item where Remote controlled compression cone is specified and they received 3 compliant bids . All firms can give Remote controlled compression cone in their high end models. You may get made in china poor quality system if you do not specify Remote controlled compression cone.</p>	<p>Remote controlled compression device/ cone for G.I studies is preferable.</p>

Item No.5 Digital Mobile X- Ray		
Tender Specifications Item No.5. Digital Mobile x-Ray	Bidder representation	Amended as
Out of three major components (Detector, X-ray tube & X-ray Generator) at least 2 must be from the same manufacturer .	<p>A) has request: Point to be deleted</p> <p>Justification</p> <p>Perhaps this clause has been set in the specs for reasons of ensuring quality. However, this limits the participation to few manufacturers and barring others. Moreover, the European CE and USFDA certification agencies provide the certificate to the entire unit irrespective of the manufacturers of the individual components. We request you to kindly delete this point for wider participation.</p>	Deleted "..... Out of three major components (Detector, X-ray tube & X-ray Generator) at least 2 must be from the same manufacturer ".

All other terms and conditions of the tender enquiry document shall remain unchanged.
Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

**Sr. Chief General Manager-I, HSCC (I) Ltd
For & on behalf of Director CNCI, Kolkata**