

HSCC (India) Ltd

AMENDMENT No – XIV Dated 08/02/2016

Sub.: Procurement of Medical Equipment for Kalpana Chawala Government Medical College, Karnal.

Ref: Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2015/02 dt. 06.10.2015

Bidder is requested to note the following technical Amendment:

Item No. 45 – MRI 1.5 Tesla

S. No.	Tender Specification	Amended Specification
1	1. MAGNET c. It should have at least 70 cm patient bore with flared opening.	1. MAGNET c. It should have at least 60 cm patient bore with flared opening.
2	1. GRADIENT SYSTEM b. The gradient should be actively shielded with each axis having independently a slew rate of at least 200 T/m/s and a peak amplitude of 33 mT/m.	2. GRADIENT SYSTEM b. The gradient should be actively shielded with each axis having independently a slew rate of at least 120 T/m/s and a peak amplitude of 33 mT/m.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

**CGM, HSCC India Limited
For and on behalf of DGMR, Panchkula**