

Amendment-XIV

Ref: IFB No. HSCC/SJH/MEDICAL EQUIPMENT/2016/31 Dated 23.06.2017

Subject : Procurement of Medical Equipment for New Emergency Block & Super Specialty Block at Safdarjung Hospital.

Regarding item no. 1, 3 Tesla MRI amendment from the Safdarjung Hospital has been received the detail is given below. Therefore It is proposed to the bid submission date is extended from 22.09.2017 to 26.09.2017. No further amendments to consider.

Existing As

Point 8. Coil System:

Point: Vi Bilateral Breast Coil with at least 16 channels with fully functional spectroscopy. Separate 4 or more channel laterally open breast coil should be provided (specify number of channels).

To be read as :

Point 8. Coil System:

Point: Vi Bilateral Breast Coil with at least 16 channels with fully functional spectroscopy. Separate 4 or more channel laterally open breast coil should be provided for biopsy.

Amendment to be issued will be uploaded on websites www.tenderwizard.com/HSCC & www.hsccltd.com.

All other tender terms and conditions remain unchanged

Medical Superintendent
VMMC & Safdarjung Hospital
New Delhi