

## AMENDMENT No.XV–Dated 24/4/2018

**Sub.: Procurement of Medical Equipment for Kalpana Chawala Government Medical College, Karnal.**

**Tender Enquiry No.:** HSCC/KCGMC/Medical Equipment/2017-06/16 uploaded on 14.12.17

**Bid sale, submission and opening date of item No. 4, 5 has been extended as per details given in Table -1:**

**Table -1**

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	01.05.18, 3.00 P.M.
ii.	Closing date & time for receipt of tender	01.05.18, 3.30 P.M.
iii.	Time and date of Opening of Tenders	01.05.18, 4.00 P.M.

Bidders are requested to note the following technical Amendment:

Item No. 4      Microlaryngoscope

S. No.	Tender Specification	Amended as
4	Adjustable Laser laryngo-pharyngoscope with integrated smoke aspiration, including light rod and suction tube length 180mm or more mat finish	The said instrument mentioned at S. No. 4,7,8,22,23 of similar technical specification, should be provided of same Manufacturer (OEM) or any other make of US FDA approve firm.
7	Grasping forceps, serrated curved right insulated working length 220 mat finish	
8	Grasping forceps, serrated curved left insulated working length 220 mat finish	
22	Clip applicator / forceps 5 mm diameter mat finish, working length 230mm	
23	Package of clips, LT 200 mm, sterile, for use with above cliip applicator	

Fiber Optic Bronchoscope

S. No.	Tender Specification	Amended as
<b>A</b>	<b>Video Processor with Built in LED Light source</b>	
	It should be integrated with 150 or more Watt LED light source with color temperature of around 6000K and a LED lamp as Auxiliary backup	Integrated Xenon /LED light source with color temperature of around 6000K and a LED lamp as Auxiliary backup
	It should have high definition video output DVI-D(2), SXGA with HD resolutions	It should have high definition video output DVI-D/HD-SDI with HD resolutions

S. No.	Tender Specification	Amended as
	It should have external device control for printer control connector with 3 configurable external device control connector	Facility for having connectivity with external device like printer etc.
<b>B</b>	<b>Adult Bronchovideoscope</b>	
	Total length of around 900mm	Total length of around 800 - 900mm
	Distal end diameter 6.8mm	Distal end diameter 5mm - 6.8mm
	Insertion tube diameter 6.4 mm	Insertion tube diameter 4.8 - 6.4 mm
	Scope PVE connector should be rotatable by 180 degree	Scope PVE connector should be rotatable by 180 degree (optional function)
	Additional Amendment required	
	Working length of 600 mm	Working length of 500-600 mm
	Field of view of view 120 degree	Field of view of view 100-120 degree
	Bending Range (deg) up 180 , down 130	Bending Range (deg) up 180 , down 100-130
<b>C</b>	<b>Pediatric BronchoVideoscope</b>	
	Rigid Distal width 3.5mm or less	Rigid Distal width 3-4mm
	Minimum Instrument channel width of 1.2mm or more	Minimum Instrument channel width of 1.5-2mm
	<b>Additional Amendment required</b>	
	Field of view of view 120 degree or more	Field of view of view 90-100 degree
	Tip Deflection up down of up 180 degree or more , down 130 degree or more	Tip Deflection up down of up 180 degree or more , down 100-130 degree or more
	Working length of 600 mm or more	Working length of 500-600 mm
	Total length of (mm) 860 or more	Total length of (mm) 800-900 mm

<b>S. No.</b>	<b>Tender Specification</b>	<b>Amended as</b>
<b>D.</b>	<b>HD Monitor</b>	
	24" or more HD, Sony Medical Grade Monitor should be provide with the system	24" or more HD, Monitor of Medical Grade or of same OEM should be provide with the system
	Should have DVI port	Should have DVI/HD-SDI port
<b>E.</b>	<b>HD Recording System</b>	<b>HD Recording System form the same make</b>
	Should have at least 4.5" screen display	Should have at least 4.5" or above screen display
	Should be FDA class 1 device , should have safety certificate like IEC60601-1, CSA 60601-1, EN60601-1	Should be FDA class 1 device , should have safety certificate like IEC60601-1, CSA 60601-1, EN60601-1 or US FDA approved
	F. Mobile Trolley	Trolley to be from the same OEM

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

**S/d**  
**Sr. CGM, HSCC India Limited**  
**For and on behalf of DGMER, Panchkula**