

HSCC (I) Ltd.

Amendment – XV

dated 2.1.2018

**Tender Enquiry No. : HSCC/KCGMC/Medical Equipment/2015-06/15
dated: 15/09/2017**

Bid sale, submission and opening date for all items has been extended as per details given (Except the items No.01,03,9,10,11,12,13,15,18) in Table

Sl. No.	Description	Revised Schedule
1	Sale date of the tender	09.01.18, 3.00 P.M.
2	Closing date & time for receipt of	09.01.18, 3.30 P.M.
3	Time and date of Opening of	09.01.18, 4.00 P.M.

Bidders are requested to note the following technical Amendment in the item no. 2

S. No.	Tender Specification	Amended as
Point No. 1 X ray Generator -C	Output Power 50 KW or more	Output Power 45 KW or more
Point No(II) X ray Tube -C	One pair of high tension cable(at least 8 meters)	One pair of high tension cable(at least 6 meters)
Point No. (VI) Others -b	QA test should be done free of cost during warranty period	First QA test should be done free of cost soon after the installation
Point No (VIII) Power Supply and UPS requirement - C	High Voltage filament transformer with silicon rectifier solenoid operated change over switch	Deleted

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

s/d

**CGM, HSCC India Limited
For and on behalf of DGMR, Panchkula**