

HSCC (I) Ltd.

Amendment – XV

dated 8.1.2018

Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2016-17/MC-06 Dated 01/09/2017

Bidders are requested to note the following technical Amendment in the specification of item No. 10

Item No. 10 GROSSING STATION

- Structure should be made completely of highly - polished corrosion Resistant stainless steel (AISI 316),material thickness 1.5 mm (table plate 2.0 mm).
- Should have Dimension Minimum L/W/H; 2000 X 1000 X 1900 mm
- Ergonomic knee space adjustments for seated working height
- Should have Height adjustable by hydraulic cylinders from 750 – 1050 mm, programmable for minimum 4 users
- Should have Downdraft and rear extraction
- Exhaust duct for connection to building ventilation system with adjustable bellows
- Should have minimum 4 LED lamps with dimmer switch
- Extracted working area should be minimum (900 X 800 mm) with 2-3 non perforated removable plates
- Dissecting area rinse from three sides (3 SIDED PERIMETER RINSE) to provide flow of water below the work area with control
- Drainage system, circulating cold water underneath the work area, with cut off valve
- Basin should be minimum (400 X 400 X200 mm),sound attenuating insulated, seamless welded, free from sharp edges and well-polished, with overflow
- Mixing tap for cold and warm water, with elbow lever, extendable shower hose (1.5 m) with shower head
- Magnetic Tool Bar, Glove dispenser, PE cutting board with steel measure
- Formalin waste drain (85 X 85mm) covered by a fine screen, Fresh formalin dispenser 10 l. Tank for waste formalin 10
- Ventilated waste container with sliding lid on table plate, Service door and telescopic shelf for formalin tanks
- Electrical control box with separate fuse for all the electrical equipment, Power supply, Sockets, main sockets as per need according to voltage/ampereage/frequency
- Footswitch for formalin pump, Cupboard and adjustable shelf of minimum dimensions 40.6 x 20.3cm,form holder, Tube holder
- MONITER AND KEYBOARD MOUNT for full articulation of the flat panel monitor and keyboard, should folds up right over mount and extends up to 11”
- MICROPHONE ON FLEX ARM should be Easy-to-adjust, small-diameter, alternating gooseneck to provide a steep low frequency attenuation to improve sound pickup without affecting voice quality with digital voice recorder
- Should have ARTICULATING CAMERA MOUNT with a CCD camera having 20x”optical zoom and at least 12 Megapixels which should have an articulated arm with large locking knob with variable friction. Should be supplied with a quick release camera plate and secondary safety lock
- Should have STAINLESS STEEL STORAGE CONTAINERS UTILITY DRAWER and PULL OUT WRITING PLATFORM if dimension not less than 60cm x 60 cm.
- Fume hood with filter system
- Should provide 5 years warranty and 5 Years CMC
- Bidders should have Service track record for at-least 10 years
- Should quote latest model as per above specification

- Demonstration is mandatory
- Compliance report to be submitted in a tabulated and point wise manner clearly mentioning the page and para of original catalogue
- List of important spare parts and accessories should be available for next 10 Years
- Installation and commissioning of equipment to be done at suppliers own cost
- Certification – ISO and CE certified and should meets international quality standard
- Dedicated USB Parts for camera control and data transfer adjustable
- Height adjustable stainless steel chairs with Split AC of appropriate capacity
- Electronic weighing scale capacity 5000 gm
- Measuring scale in cm & inches etched or fixed on the work surface
- To work at 200 -240 V
- Small specimen rinse with cold water control valve & sink rinse basket with stainless steel holder
- Document supporting track record and satisfactory performance from institution of national importance (minimum of one) should be provided.

Bid sale, submission and opening date for items 10 has been extended as per details given in Table -1:

Sl. No.	Description	Revised Schedule
1	Sale date of the tender	16.01.18, 3.00 P.M.
2	Closing date & time for receipt of tender	16.01.18, 3.30 P.M.
3	Time and date of Opening of Tenders	16.01.18, 4.00 P.M.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

s/d
CGM, HSCC India Limited
For and on behalf of DGMER, Panchkula