

AMENDMENT No. XVIII – Dated 21/05/2018

Sub.: Procurement of Medical Equipment for Kalpana Chawala Government Medical College, Karnal.

Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2017-06/16 uploaded on 14.12.17

Bid sale, submission and opening date of item No. 4, 5 have been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	28.05.18, 3.00 P.M.
ii.	Closing date & time for receipt of tender	28.05.18, 3.30 P.M.
iii.	Time and date of Opening of Tenders	28.05.18, 4.00 P.M.

Bidders are requested to note the following technical Amendment:

Item No. 4 – Microlaryngoscope

S. No.	Description	AMENDED/ READ AS
4	Adjustable Laser laryngo-pharyngoscope with integrated smoke aspiration, including light rod and suction tube length 180mm or more mat finish	Should be provided of same Manufacturer (OEM) or any other make of US FDA approved firm.
7	Grasping forceps, serrated curved right insulated working length 220 mat finish	Should be provided of same Manufacturer (OEM) or any other make of US FDA approved firm.
8	Grasping forceps, serrated curved left insulated working length 220 mat finish	Should be provided of same Manufacturer (OEM) or any other make of US FDA approved firm.
22	Clip applicator / forceps 5 mm diameter mat finish, working length 230mm	It can be US FDA or European CE from other make
23	Package of clips, LT 200 mm, sterile, for use with above clip applicator	It can be US FDA or European CE from other make

Item No. 5 Flexible Video Bronchoscope

S.No.	Description	AMENDED/ READ AS
	Flexible Video Bronchoscopy System for ENT Dept.	
	Should consists of following components:	
A	Video Processor with built in 150 watt Xenon light source	
	It should have high definition video output DVI-D/HDS DI with HD resolutions	“All the output should ensure progressive scan images and videos” to original technical specification
E	HD Recording System from the same make	HD recording system from same manufacturer (OEM) or any other make of US FDA approved firm.
F	Trolley to be from the same OEM	Trolley to be from the same Manufacturer

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

S/d
CGM, HSCC India Limited
For and on behalf of DGMER, Panchkula