

All Bidders

Amendment-1

Subject: **Tender for** Package-II : Supply, installation, testing & commissioning of Hospital related Furniture & other associated works like minor civil, electrical, PHE works etc. for Emergency Block at Safdarjung Hospital & VMMC, New Delhi

Dear Sir,

This has reference to the captioned subject works,

The following AMMENDMENT-1/CLARIFICATION-1 may be noted, which shall be treated as a part of tender **to be submitted online duly signed & stamp along with tender**

BOQ item no	Description (existing specification)	Clarifications (Modified soecification)
03	<p>ICU BED with Mattress :</p> <p>Fully automatic wire remote control with back rest up down, knee rest up-down, trendelenburg and reverse trendelenburg, hi-low position controlled through noiseless electro mechanical actuators (Linak made) operated by soft touch attendant(nurses')control panel.</p> <p>The bed should be provided with 125mm plastic injection moulded twin wheel castors with central and directional locking facility (Tente make). The castors should be provided with MS round tube made auxiliary brake.</p>	<p>Fully automatic wire remote control with back rest up down, knee rest up-down, trendelenburg and reverse trendelenburg, hi-low position controlled through noiseless electro mechanical actuators (TI Motion) operated by soft touch attendant (nurses') control panel.</p> <p>The bed should be provided with 125mm plastic injection moulded single wheel castors with diagonal locking facility (Tente make).</p>
16	<p>Examination Couch :</p> <p>Top Drawer Unit is provided, made of CRCA sheet construction with SS handle and SS hinge. Sheet thickness 1.2 ,0.8 and</p>	<p>BP Apparatus tray: CRCA sheet construction with SS handle and SS hinge</p>

	0.6. Two drawers at top; each of volume 510 x 334 x 95 mm are provided	
6	Crash cart: Overall dimensions L 855 W 405 H 1520 mm (These are internal dimensions considered)	To be changed to L 1048 W 475 H 1555 (These are external dimensions)]
10	Instrument trolley with Bowl & Bucket: Dimension must be 1232 mm X 531 mm X 915 mm (Last Line)	Dimension must be 1156 mm X 531 mm X 915 mm (Last Line)

Other all terms & conditions of tender shall remain unchanged.

Prospective bidders are advised to regularly scan through HSCC e-tender portal <http://www.tenderwizard.com/HSCC> as corrigendum/amendments etc, if any, will be notified on this e- tender portal only and separate advertisement will not be made for this.

- sd-
(A.K.Sinha)
Project Coordinator