

**HSCC (I) Ltd**

**Amendment - XX Dated 13.01.2021**

**Ref. IFB No: HSCC/AIIMS-RAEBARELI/Hospital/07/2020 Dated : 04.09.2020**

**Extension of Bid Submission Date**

**Technical Amendment:**

Name of Item& No.: **18, 532 nm Green Laser**

Department: Ophthalmology

<b>Specification Sr. No.</b>	<b>Tendered Specification</b>	<b>To be Read As</b>
7	SLIT LAMP – Good Quality, Reputed Laser Slit Lamp With 5-Steps Magnifications Halogen Illumination, Continuously Adjustable Light Intensity; Integrated Joystick Controlled Servo-Electric Micromanipulator and True-To-Colour Physician’s Safety Filter Automatically Swinge-able into position; Slit height adjustable from 1 to 12mm or more and slit width adjustable from 1 to 12mm or more	Good Quality, Reputed Laser Slit Lamp With 5-Steps Magnifications Halogen/LED Illumination, Continuously Adjustable Light Intensity; Integrated Joystick Controlled Servo-Electric Micromanipulator or Manual Micromanipulator and True-To-Colour Physician’s Safety Filter Automatically Swinge-able into position; Slit height adjustable from 1 to 12mm or more and slit width adjustable from 1 to 12mm or more.
13	Halogen Lamp	Halogen/LED Lamp

The bid submission date remains 20.01.2021.

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPPP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

**CGM (PROC.)  
HSCC (I) LTD.**