## Amendment -XXI

dated 18.12.2017

Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2015-06/14 dated: 30/06/2017

Bid sale, submission and opening date for items 05 has been extended as per details given in Table -1:

Table -1

SI. No.	Description	Revised Schedule
i.	Sale date of the tender	25.12.2017, 3.00 P.M.
ii.	Closing date & time for receipt of tender	25.12.2017, 3.30 P.M.
iii.	Time and date of Opening of Tenders	25.12.2017, 4.00 P.M.

Bidders are requested to note the following technical Amendment in the item no. 05 Battery Operated Drill

Point No.	Specification as per the tender	Amended as
7	Batteries-04	
	Long battery lifetime (>1000 charging cycles)	its 200 or more charging cycle
9	Battery Charger-01	
	Should have 4 charging bays	it should be from 2-4 charging bay

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC e tender website/ CPPP websites <a href="https://www.tenderwizard.com/HSCC">www.tenderwizard.com/HSCC</a> & <a href="https://www.tenderwizard.com/HSCC">www.hsccltd.co.in</a>, <a href="https://www.eprocure.gov.in">www.eprocure.gov.in</a> for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

s/d CGM, HSCC India Limited For and on behalf of DGMER, Panchkula