


	<p>inlet vacuum level inlet 578mm Hg to 728mm Hg is company specific. The Bacteria filtration states that the drain flask should be made of transparent pyrex*. Pyrex is again a company specific material used by some particular company. The star mark reflects that it is a patented product. The certification is also specific as stated above.</p> <p><b>f) Master &amp; other alarm system</b></p> <p>The complete specifications and certification are company specific as explained above</p> <p><b>g) Outlets</b></p> <p>The outlets specifications and details are also company specific. It is not only the details which are company specific but even the proprietary brand of the company specific GEM10 is stated in the tender documents</p> <p><b>h) Other MGPS Equipments</b></p> <p>The complete specifications and certification are company specific as explained above</p>	
12.	<p>Third Party Certification</p> <p>As per Tender Documents every bidder need to furnish a third party certification that all the MGPS equipments as per specifications and BOQ.</p>	As per SI. No.-8
13.	<p>The bidder should have installation of the same foreign products in India and it should be functional for at least last one year without any faults and complaints. A certificate to be submitted from the customer.</p>	Tender terms prevail
14.	<p>CMC should be provided by for foreign principal company. So that incase on Agency of Indian agent is terminated the hospital does not suffer.</p>	Tender terms prevail
15.	<p>Please clarify who will do construction of Gas Manifold and Plant room</p>	As per SI No-2
16.	<p>Who will handle the A.C part of plant room ??</p>	Tender terms prevail

  
 12/9/14

 4  
 12/9/2014

17.	Country of origin certificate must be submitted.	Country of origin certificate must be submitted by the bidder.
18.	Company should have trained engineers from foreign company whose products are being quoted copy should be submitted.	Tender terms prevail
19.	Company who will be doing the installation should have ISO14001 certificate from notified body.	Tender terms prevail
20	Pendant	Tender terms prevail
21	Standards	
	All imported items should be international standard like NFPA 99(latest version) standard and UL listed /CE marked or HTM02-01 (latest version) standard and CE marked.	As per Sl. No.-11
22	The entire system should be of anyone standard only.	As per Sl. No.-11
23	All indigenous items should be of highest quality to meet the international standard and compatible to the main system.	As per Sl. No.-11
	Page 36	
24	Fully Automatic Oxygen Control Panel	As per Sl. No.-11
	NEMA Enclosure	As per Sl.No.-11
	Switch should be mechanically controlled, not electrical.	As per Sl. No.-11
	Page 38	
25	Fully Automatic Nitrous Oxide Control Panel	As per Sl. No.-11
	NEMA Enclosure	As per Sl. No.-11
	Page 39	

  
 12/9/14  
 12/9/2014  
 5