

HSCC (I) Ltd

Amendment –XLIV Dated 17.11.2020

IFB Ref No :HSCC/AIIMS-RAEBARELI/Hospital/2/2019 Dated : 28.11.2019

Technical Amendments Gyne & Obst. Department are as follows:

Item No. 2, Portable Ultrasound

Department: Gyn.&Obs.

Specification Sr. No.	Tendered Specification	To be read as
2	The Unit should have a laptop type console design the unit should be compact lightweight and portable. Weight should not less than or equal to 11 kg including battery (Excluding cart and accessories)	The unit should be compact lightweight and portable weight should not exceed 5 kg. including battery (excluding cart and accessories)
8(2)	Linear transducers 5-12 MHz for vascular and small part imaging.	Linear Transducers 5-12 MHz (+/-1 MHz) for vascular and small part imaging.
14	The system should have the facility of digital storage and retrieval of B/w and color image data on built in CD/DVD drive. Provision for USB port and LAN transfer of data should also be presence .	The system should have the facility of digital storage and retrieval of B/w and color image data on built in flash memory card /HDD. Provision for USB port.
21	Facility for storage on CDR should be available	Deleted
23	Inbuilt battery backup should be at least one hours or more	Inbuilt battery backup should be at least two hours or more
	Point Added	Should be US-FDA or European CE with 4 Digit notified body no. or BIS

The bid submission date of all above items have been extended from **17.11.2020 to 24.11.2020**

Further Amendments of the following Equipment are pending; their the bid submission date is extended from **17.11.2020 to 24.11.2020**.

1. (Orthopedics Department Items No. B1, B2, B3, B4, B5),
- 2.. (CTVS Departments Item No. D3, D4)
3. (Neurosurgery Department Items No. E2,E3,E5,E6)
4. (Blood Bank Items No. L13 L15,)
5. (PMR Department Item No. M1,M2)
6. Item No. N8 (C-Arm with IITV)
7. Item No. N6D (OT Tables for different Departments)

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPMP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

**CGM (PROC.)
HSCC (I) LTD.**