

HSCC (I) Ltd

Amendment –XLIX Dated 23.12.2020

Ref. IFB No: HSCC/AIIMS-RAEBARELI/Hospital/02/2019 Dated : 28.11.2019

Subject: Procurement of Medical Equipment for AIIMS -Raebareli.

Technical Amendment:

Name of Item& No.: **B1, Orthopedic Table with Attachment**

Department: Orthopedics Department

Specification Sr. No.	Tendered Specification	To be Read As
1	Radiolucent table top made up of Carbon fibre or equivalent for orthopaedic use	Radiolucent table top for orthopaedic use
3	Should have position for eccentric position	Should have provision for Eccentric/Centric provision
4	Two foldable and detachable radiolucent traction bars fixed beneath the seat plate with two adjustable pivot joints	Two foldable and detachable radiolucent traction bars fixed beneath the seat plate & two adjustable pivot joints
5	Accessories Traction bards radiolucent – 02	Accessories: Traction Bars – 02
7	Additional radiolucent attachment/ plate for orthopedic intervention for pediatric patients	-DELETED-
09A	Table top height range – 70cm to 120cm	Table top height range – 70cm to 112cm
09F	Battery capacity for nearly 2 weeks with average use	Batteries to hold charge up to 1 week or up to 80 operations
09J	Length: 210-220cm	Length: 205cm to 220cm
13	Accessories for genucubital position/ Accessories for genupectorial position	Accessories for genucubital position
15	Mayfield attachment for cervical spine	Mayfield attachment for cervical spine/ DORO optional
17	Accessories for interlocking nailing for femure in supine position	-DELETED-
19F	Gel pads for different positions	-DELETED-
20	Cushions(one set each): as foam pads for different positions: Head Ring, Lateral positioning, Leg rest cushion, cushions especially for spine surgery	-DELETED-
21	Should include below Pediatric accessories: a) Paediatric traction boot b) Side Post c) Perineal Post	-DELETED-
22	Two or more detachable shoulder segment	-DELETED-
23	New Point Added:	Should be US-FDA or European CE with 4 Digit notified body number or BIS

All other Specifications shall remain unchanged.

Name of Item& No.: **B2, Automatic Tourniquet System**

Department: Orthopedics Department

Specification Sr. No.	Tendered Specification	To be Read As
	New Point Added:	Should be US-FDA or European CE with 4 Digit notified body number or BIS

All other Specifications shall remain unchanged.

Name of Item& No.: **B3, Battery Operated Drill/ Reamer and Saw for Joint Replacement**

Department: Orthopedics Department

Specification Sr. No.	Tendered Specification	To be Read As
1 Para-I	Should have Dual trigger for Forward/Reverse and Oscillation Mode	Should have Dual trigger for Forward/Reverse
1 Para-II	Minimum speed of 1200 rpm and should have variable speed control on the Handpiece	Minimum speed of 900-1200 rpm and should have variable speed control on the Handpiece
1 Para-III	Drill torque 30 in lbs or more	Drill torque 15-30 in lbs or more
7 Para-I	Should have two speed control with Standard & Fast Mode	Should have two speed control with Standard & Fast Mode (or) variable speed Slow- Standard- Fast control is achieved by pressing a single trigger
7 Para-II	Blade Mount should be adjustable to different angles movement adjustable in 8 steps to 45 degree	Blade Mount should be adjustable to different angles movement adjustable in 8-12 steps to 30-45 degree
12	Metal Cutting	-Deleted-
13	Burrs for Cement removal and sculpting	-Deleted-
14 Para-III	Should have Nickel Cadmium: autoclavable standard and Small Battery	Should have Nickel Cadmium/ Lithium ion autoclavable standard and Small Battery
	New Point Added:	Should be US-FDA or European CE with 4 Digit notified body number or BIS

All other Specifications shall remain unchanged.

Name of Item& No.: **B4, General Orthopedic Instruments**

Department: Orthopedics Department

Specification Sr. No.	Tendered Specification	To be Read As
32 Para-II	-Deleted-	Should be US-FDA or European CE with 4 Digit notified body number or BIS
	New Para Added :	Variation in size upto 10% is acceptable. Bidder should quote all the instruments and 80% of Instruments should be from the same manufacturer Balance 20% must be from Other manufacturer with same certification
15 Para-II	-Deleted-	Should be US-FDA or European CE with 4 Digit notified body number or BIS
	New Para Added :	Variation in size upto 10% is acceptable. Bidder should quote all the instruments and 80% of Instruments should be from the same manufacturer Balance 20% must be from Other manufacturer with same certification
19 Para-II	Instrument should be European CE/USFDA approved	Should be US-FDA or European CE with 4 Digit notified body number or BIS
	New Para Added :	Variation in size upto 10% is acceptable. Bidder should quote all the instruments and 80% of Instruments should be from the same manufacturer Balance 20% must be from Other manufacturer with same certification
12 Para-II	Instrument should be European CE/USFDA approved	Should be US-FDA or European CE with 4 Digit notified body number or BIS
	New Para Added :	Variation in size upto 10% is acceptable. Bidder should quote all the instruments and 80% of Instruments should

		be from the same manufacturer Balance 20% must be from Other manufacturer with same certification
4 Para-V	Deleted	Should be US-FDA or European CE with 4 Digit notified body number or BIS
	New Para Added :	Variation in size upto 10% is acceptable. Bidder should quote all the instruments and 80% of Instruments should be from the same manufacturer Balance 20% must be from Other manufacturer with same certification
1-Needle Holder Para-V	Deleted	Should be US-FDA or European CE with 4 Digit notified body number or BIS
	New Para Added :	Variation in size upto 10% is acceptable. Bidder should quote all the instruments and 80% of Instruments should be from the same manufacturer Balance 20% must be from Other manufacturer with same certification
All Specifications	All the sub sets are under this category have to be quoted together	All the sub sets under this category i.e. : Set no: 1,Set no: 2, Set no: 3, Set no: 4, Set no: 5, Set no: 6 should be quoted sseperately
	New Point Added:	Demonstration, if needed has to be given
	New Point Added:	Quoting company should have relevant experience in successful execution of similar work in at least 4 to 5 institutes of National Importance for Central Government Institutes. Performance certificate of at least 2 institutes of the same should be attached.
	New Point Added:	Should be US-FDA or European CE with 4 Digit notified body number or BIS

All other Specifications shall remain unchanged.

Name of Item& No.: **B5, Arthroscopy System**

Department: Orthopedics Department

Specification Sr. No.	Tendered Specification	To be Read As
3	Arthroscopy Shoulder Set and Arthroscopy Ankle Surgery Set imaging system	-DELETED-
6	Xenon Light Source : High Definition Monitor 19"	Xenon Light Source : Medical Grade Monitor 26" and Above
	New Point Added :	The System Should be USFDA Approved /European CE with Four digit notified body number /BIS

All other Specifications shall remain unchanged.

Name of Item& No.: **M1, Operation theater instruments**

Department: Physical Medicine and Rehabilitation (PMR) Department

Specification Sr. No.	Tendered Specification	To be Read As
	New Point Added :	The System Should be USFDA Approved /European CE with Four digit notified body number /BIS
	New Point Added :	Price of individual instruments should be quoted
	New Point Added :	Demonstration if needed has to be given
	New Point Added :	Instruments should be made from High Quality Surgical

		Grade Steel i.e. Grade 410 & 420 or equivalent
	New Point Added :	Instruments should have laser surface or ebonized or equivalent finish to provide appropriate reflection lowering finish
	New Point Added:	Catalogue number & Article Number should be mentioned in each and every instrument
	New Point Added:	There should be Country of Origin/ Manufacturing engraved on each and every instrument
	New Point Added:	Certificate of Origin to be issued by Accredited Chamber of Commerce of the manufacturing country
	New Point Added:	Quoting Company should have relevant experience in successful execution of similar work in at least 4 to 5 Institutes of National Importance or Central Government Institutes. Performance Certificate of the same should be attached.
	New Point Added:	Manufacturer should have their own repair center in India

All other Specifications shall remain unchanged.

Name of Item& No.: **M2, Tilted Table (Motorized)**

Department: Physical Medicine and Rehabilitation (PMR) Department

Specification Sr. No.	Tendered Specification	To be Read As
	New Point Added :	The System Should be USFDA Approved /European CE with Four digit notified body number /BIS

All other Specifications shall remain unchanged.

Name of Item& No.: **E2, Neuro Surgical Drill with all attachments**

Department: Neurosurgery Department

All Specifications shall remain unchanged.

Name of Item& No.: **E3, Intra Operative Nerve Monitoring**

Department: Neurosurgery Department

All Specifications shall remain unchanged.

Name of Item& No.: **E5, Neuro Surgical Instrument Set**

Department: Neurosurgery Department

All Specifications shall remain unchanged.

Name of Item& No.: **E6, Operating Microscope**

Department: Neurosurgery Department

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPPP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

**CGM (PROC.)
HSCC (I) LTD.**