

Amendment No. 50 Dated 24.12.2018
HSCC/PUR/CNCI/Kolkata/Medical Equipment/03 dt. 15.11.2017

Procurement of Medical Equipment CNCI 2nd Campus, all bidders are requested to note the following for “CT Scan 128 Slice” technical specification:-

Item No	Name of the equipment / Instrument	Last date & time sale/downloading of Tender document	Last date & time closing/submission for receipt of tender	Date of opening of Techno – Commercial bids.	Earlier date of opening	Amendment Status for opening date
2	CT Scan 128 Slice	18.01.2019, up to 13.00 hrs IST	18.01.2019, 14.00 hrs IST	18.01.2019, 14.30 hrs IST	03.01.2018	Date Extended

Tender Specifications	May please read as
1a. High frequency generator with output of at least 70 KW	Tender terms prevails.
2a. Anode heat storage capacity of 7.0 MHU or higher or the tube with direct cooling technology	Tender terms prevails
7b. The scan time for a 360 Degree rotation should be 0.35 second or lower.	The scan time for a 360 Degree rotation should be 0.4 second or lower
8c. Latest Modality based Iterative Reconstruction Technique to be quoted as standard (Admire/Asir-V/IMR/or equivalent). Certified document showing extent of dose reduction by these techniques should be attached	Latest Modality based Iterative Reconstruction Technique to be quoted as standard (Admire/ASIR -V/IMR/MBIR First/ AIDR-3D or equivalent/). Certified document showing extent of dose reduction by these techniques should be attached
16.xiii.Tumour ablation system with treatment planning solution & RF generator.	Deleted
16.iv. Patient Positioning Accessories: Head Rest, Head and Arm Support, Knee and Leg Support, Coronal Supine Head Holder. Paediatric Immobilizer.	Patient Positioning Accessories: Head Rest, Head and Arm Support, Knee and Leg Support, Coronal Supine Head Holder. Paediatric Immobilizer. The material of all Patient Positioning accessories should be Carbon fibre.
16.V.This accessories added	One Indexed Flat table top made of Carbon fibre compatible with Linac(Elekta-Versa-HD/ Varian-True beam)

All other terms and conditions of the tender enquiry document shall remain unchanged. Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum / amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the newspapers.

Sr. Chief General Manager-I, HSCC (I) Ltd
For & on behalf of Director CNCI, Kolkata