

Amendment no V Dated 21.03.2017
HSCC/KCGMC/Medical Equipment/2015-03/G Dated: 03.02.2017

All bidders are requested to note the following:

Item	Tender Specifications	Tender Specifications may please read as
Item no 01, Refrigerated Centrifuge	3. Swing bucket blood bank rotor:- With metal buckets, 6 x 1000 ml. wind-shielded. Suitable adapters (Removable plastic oval 1 cups) for 12 blood bags of 350 ml. & 450 ml. -2 sets each. Insert with hook adapter to spin buffy coat or small volumes of blood. Balancing weights for inserts.	3.“SWING OUT ROTOR WITH OR WITH-OUT WIND SHIELD”
	10 Easy to clean, corrosion resistant with provision of both drain and condensed water collection container. 10. Stainless steel chamber:- Easy to clean, corrosion resistant with provision of both drain and condensed water collection container.	10 “Stainless steel chamber: easy to clean and corrosion resistant”.
	11. Power requirement 11220/240 volts, 50 Hz. AC supply.	11. “Power requirement to be 200- 400V, single or three phase, as per machine requirement”.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

s/d
CGM, HSCC (I) Ltd.
For & on behalf of DGMER, Panchkula