

# HSCC (India) Ltd

---

## AMENDMENT – VIII

Dated: - 16.09.2016

**Subject:** Amendment to the tender Enquiry Document.

**Ref: Tender Enquiry No.:** HSCC/KCGMC/ Medical Equipment / 2015-06/04 Dated 25.06.2016

**Bid sale, submission and opening date for item no. 2,3,4,6,7,8 has been extended as per details given in Table -1:**

**Table -1**

<b>Sl. No.</b>	<b>Description</b>	<b>Revised Schedule</b>
i.	Sale date of the tender	23.09.2016, 2.00 P.M.
ii.	Closing date & time for receipt of tender	23.09.2016, 2.30 P.M.
iii.	Time and date of Opening of Techno – Commercial Tenders	23.09.2016, 3.00 P.M.

Bidders are requested to note the following technical Amendment

### **Item no. 2 Flexible Sigmoidoscope**

A	<b>Video Processor with inbuilt light source</b>	<b>Amended to</b>
2	It should have single CCD color system	It should have 3CCD /HD color system
3	It should have 150watts Halogen lamp or more	It should be Xenon Light source ( 100W or more)
B	<b>Video Sigmoidoscope</b>	
9	It should have a rotatable PVE Connector for easier manoeuvrability by 180 Degree or equivalent	It should have a rotatable PVE Connector for easier manoeuvrability by 180 Degree or equivalent
		Indigenous Make Trolley for Camera and Light Source

## HSCC (India) Ltd

### Item no. 4 C Arm

Tender Spec No.	Tender Specification	Amended to
3 Technical specs 1.General - Point a	Motorized Vertical travel : MINIMUM 430 mm or more	Motorized Vertical travel : MINIMUM 420 mm or more
point b	Pivotal rotation : =/- 12.5 deg. Or more	Pivotal rotation : =/- 12.0 deg. Or more
5. X- ray generator	High frequency 3 KW or even better X-ray generator with high capacity rotating anode X-ray tube of dual foci	High frequency 2.5 KW or even better X-ray generator with high capacity rotating anode X-ray tube of dual foci
6. a	Fluoroscopy output : 40-120 KV in IKV steps	Fluoroscopy output : 40-110 KV in IKV steps
7.a	Radiography output : 40-120 KV in 1 KV steps	Radiography output : 40-110 KV in 1 KV steps

### Item No. 6 Suction Machine

Description	Amended to
Should have following facility. -	
Fast vacuum build up, vacuum capacity min. 40L/min, min. -700mm Hg	Fast vacuum build up, vacuum capacity min. 30L/min, min. -700mm Hg
2lt.x2lt. double suction container, PC/polysulfone, graduated with lid for overflow protection and 2 spare suction tubes.	3lt.x3lt. double suction container, PC/polysulfone, graduated with lid for overflow protection and 2 spare suction tubes.

### Item No. 7 Microlaryngoscopy Set

Tender Specification	Amended to
Fibre Optic light system	Fiber optic light carrier
Microlaryngeal cup Forceps 23 cm, serrated right and left turned	Microlaryngeal cup Forceps 22-23 cm, serrated right and left turned
Microlaryngeal Scissors Straight, 23cm	Microlaryngeal Scissors Straight, 22-23cm

## HSCC (India) Ltd

---

### Item No. 8 Mastoidectomy Set

<b>Tender Specification</b>	<b>Amended to</b>
Sterilisable cutting burr & Diamond Burr from sizes	Sterilisable cutting burr & Diamond Burr from sizes  0.8 mm } 2.0 mm } 3.0 mm } 4.0 mm } 5.0 mm } 6.0 mm }  Cutting 10 (each) Diamond 5 (each)  · Irrigation accessories to be provided along with the system.
Mastoid seeker/ Gross ear hook spoon	Mastoid seeker
Aditus seeker/ gross ear hook curette	Aditus seeker

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

**CGM, HSCC India Limited**  
**For and on behalf of DGMR, Panchkula**