

All Bidders

Amendment -II

Subject: Supply, Installation, Testing & Commissioning of CSSD Equipment for PGI Satellite Centre, Sangrur, Punjab.

IFB No. : HSCC/SES/CSSD/PGI/SANGRUR/2018 dated 09.11.2018

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidder's Queries	Reply
1	<p>Page No. 3 of Volume I: Estimated Cost Exclusive CMC : Rs. 260.0 Lakhs</p> <p>Page No. 3 of Volume I: Estimated Cost Exclusive CMC: Rs. 500.0 Lakhs plus GST.</p> <p>We request you to kindly increase the estimated cost as existing estimated cost is very low and also GST should be extra.</p>	Tender terms and conditions remain unchanged.
2	<p>Page No. 5 of Volume I: Completion Period: 4 months from the date of order of commencement.</p> <p>Page No. 5 of Volume I: Completion Period: 5/6 months from the date of order of commencement.</p>	Tender terms and conditions remain unchanged.
3	<p>Page No. 3 of Volume IV: Item No. 2: Horizontal Double Door Autoclave 900-1000 Litres (14-16 STU) with Accessories</p> <p>Page No. 3 of Volume I: Item No. 2: Horizontal Double Door Autoclave 1150-1250 Litres (14-16 STU) with accessories</p>	Tender terms and conditions remain unchanged.
4	<p>Page No. 6 of Volume IV: Item No. 3: Ethylene Oxide Sterilizer</p> <p>Please add: It should be European CE Certified from notified body.</p>	Tender terms and conditions remain unchanged.

5	<p>Page No. 6 of Volume IV: Item No. 3: Ethylene Oxide Sterilizer iv) The chamber shall be insulated against heat emission and jacket shall be connected to warm water circulation arrangement.</p> <p>Page No. 6 of Volume IV: Item No. 3: Ethylene Oxide Sterilizer iv) The chamber shall be insulated against heat emission.</p> <p>Warm Water is not required in our system.</p>	Tender terms and conditions remain unchanged in the case of usage of warm water.
6	<p>vii) The sterilizer shall be provided with suitable vacuum pump and gas trap to separate and evacuate the gas.</p> <p>vii) The sterilizer shall be provided with suitable vacuum pump / blower and gas trap to separate and evacuate the gas.</p>	The sterilizer shall be provided with suitable vacuum pump/ blower and gas trap to separate & evacuate gas.
7	<p>xi) Gas cartridges should be puncturing system</p> <p>xi) Gas cartridges should have puncturing system / activating system.</p> <p>In our system cartridge is activated inside the sterilizer.</p>	Gas cartridges should have puncturing system/activating system.
8	<p>Should be provided with Compressor if required.</p> <p>Kindly delete.</p> <p>Compressor is not required in our system.</p>	Tender terms and conditions remain unchanged.
9	<p>Page No. 6 of Volume IV: Item No.. 4 : Double Door Washer Disinfector 300 Litre (10/15 DIN Trays) with Accessories</p> <p>Page No. 6 of Volume IV: Item No. 4 : Double Door Washer Disinfector 280 Litre or more (12/15 DIN Trays) with Accessories</p>	Capacity: 275-300 Litres
10	<p>Page No. 6 of Volume IV: Point No. 4 : Chamber Capacity: Volume should be 300 L. Should supply 15 Nos. of standard DIN trays.</p> <p>Page No. 6 of Volume IV: Point No. 4 : Chamber Capacity: Volume should be 280 L or more. Should supply</p>	Capacity : 275-300 Litres

	12/15 Nos. of standard DIN trays.	
11	<p>Page No. 9 of Volume IV Item No. 9 – Drying Cabinet</p> <p>It should be European CE certified (DOC) / UDFDA. Please amend the same.</p>	Tender terms and conditions remain unchanged.
12	<p>Page No. 14 of Volume IV: Item No. 20 - Modular Sterilizing Basket Size : 585x395x195 Item No. 21 – Modular Sterilizing Basket Size : 585x395x100</p> <p>Page No. 14 of Volume IV: Item No. 20 - Modular Sterilizing Basket Size : 600 x 300 x 290</p> <p>Kindly note that Specification of Modular Sterilizing Basket asked in tender is as per SPRI standard Size while Sterilizer has been asked as per STU standard. We request you to kindly amend the size as per STU.</p>	<p>Modular Sterilizing Basket size 585x395x195 / 600x300x290</p> <p>Size of Modular Sterilizing Basket shall be as per SPRI/STU Standard.</p>
13	<p>Please add: RO Plant</p> <p>RO should of Eureka Forbes /Ion Exchange / Delta Tech / Millipore / Kent / Aquacare / Rions / Water expert make</p>	<p>RO should of Eureka Forbes /Ion Exchange / Delta Tech / Millipore / Kent / Aquacare / Rions / Water expert make</p>
14	<p>Furniture Items</p> <p>1. All CSSD Furniture items should be ISO 9001.</p> <p>2. All CSSD Furniture items should be accepted with $\pm 10\%$ deviations.</p>	<p>Tender terms and conditions remain unchanged.</p> <p>All CSSD Furniture items should be accepted with $\pm 10\%$ deviations.</p>
15	<p>Horizontal Double Door Autoclave 900-1000 Litres (14-16 STU) with accessories</p> <p>As per major international models available 900 Litre autoclaves offer 12-14 STU capacity only so either it should be changed or it should be optional for either 900- 1000 Litres or 14-16 STU capacity.</p>	<p>900-1000 Litres with 14 STU± 2 STU.</p>
16	<p>Double Door Washer Disinfector 300 Litres (10/15 DIN Trays) with accessories</p> <p>The kind of door movement is missing (sliding/hinged). We also recommend to consider for 250 Litres (10 DIN Tray) which comes with Hinged Doors , keeping owning & maintenance cost quite reasonable as compared with sliding door</p>	<p>Door movement should be sliding.</p> <p>Capacity : 275-300 Litres.</p>

	models. Also 300 Litre model is costly as compared to 250 Litre although offers the same desired DIN capacity as per the tender.	
17	Ethylene Oxide Sterilizer The same should be desired for double door model for better sterilization purposes with one door opening in clean zone and the other in the sterile zone.	Tender terms and conditions remain unchanged.

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.
For and behalf of Director, PGIMER, Chandigarh