

Amendment No. VI Dated 30.10.2018
HSCC/PUR/CNCI/Kolkata/Medical Equipment/06 dt. 04.09.2018

Procurement of Medical Equipment CNCI 2nd Campus.

All bidders are requested to note the following:

Item No.01- High end C Arm

Tender Specifications	May please be read as
<p>1. Generator and X-ray Tube: A. Generator should be microprocessor controlled with power output of 15KW or more with the each of following modes :</p> <ol style="list-style-type: none">1. Continuous boost Fluoroscopy.2. Digital Pulsed Fluoroscopy.3. Digital Radiography mode. <p>B. The range of KV should be 40 to 120KV or more</p>	<p>1. Generator and X-ray Tube: A. Generator should be microprocessor controlled with power output of 20KW or more with the each of following modes :</p> <ol style="list-style-type: none">1. Continuous boost Fluoroscopy.2. Digital Pulsed Fluoroscopy.3. Digital Radiography mode. <p>B. The range of KV should be 40 to 120KV or more</p>

All other terms and conditions of the tender enquiry document shall remain unchanged. Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the newspapers.

Sr. CGM-I, HSCC (I) Ltd
For & on behalf of Director CNCI, Kolkata