INVITATION FOR QUOTATION

HSCC/PUR/Low Value/Medical Equipment – Mobile OT Light/2018 Dated: 16.02.18

To,		
	All Bidders	

Subject: Invitation for Quotations for supply of Low Value Items for setting up of Temporary OPD for AIIMS, Rae Bareli under PMSSY.

Dear Sirs,

1. HSCC (India) Ltd. for and on behalf of Ministry of Health & Family Welfare, Government of India invites your most competitive quotation for the following goods the detailed specifications are given in Annexure – B of this Invitation for quotation.

Mobile OT Light Qty -1

- 2. Quotation,
- 3.1 The contract shall be for the full quantity as described above.
- 3.2 Corrections, if any, shall be made by crossing out, initialing, dating and re writing.
- 3.3 The prices quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- 3.4 The unit price/ rate of the item should be clearly indicated in the tender. Rates /Prices quoted Inclusive of all taxes, duties, forwarding, insurance & transportation upto destination (AIIMS, Rae Bareli).
- 3. Each bidder shall submit only one quotation.
- 4. Evaluation of Quotations,

The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which are properly signed along with Make & Model must be

quoted along with literature duly marked with specification asked in quotation; and

- 6.2 confirm to the terms and conditions, and specifications.
- 5. The Quotations would be evaluated item wise.
- Award of contract:

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest rate for the item.

- 6.1 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and cancel the bidding process and reject all quotations at any time prior to the issue of Purchase order without assigning any reason.
- 7. Delivery period shall be 10 days from date of placement of order.
- 8. Payment shall be made in Indian Rupees as follows:

Satisfactory Acceptance - 100% of total cost

- 9. All supplied items shall under warranty of **12** months from the date of successful acceptance of items.
- 10. You are requested to provide your offer latest by **14:00** hours on **20.02.18** .The quotations will be opened in HSCC office at 15:00 hours on **20.02.18** in the presence of tenderer or their authorised agents such as they may choose to attend.
- 11. Information brochures/ Product catalogue, if any must be accompanied with the quotation clearly indicating the model quoted for.
- 12. Sealed quotation to be submitted/delivered at the address mentioned below,

General Manager (Procurement) HSCC (India) Ltd E- 6 (A), Sector -1 NOIDA - 201 301

THE COVER SHOULD BE SUPERSCRIBED WITH THEFOLLOWING:

Reference to letter of enquiry. Due date of opening.

- 13. We look forward to receiving your quotation and thank you for your interest in this project.
- 14. Quoted amount should be in Indian Rupees only and should be valid till 30.06.18
- 15. For all items the Technical Evaluation Committee, may opt for Demonstration of the

items. The Committee may also ask for Demonstration / Inspection before supply / delivery of the items for quality assurance.

16. Insurance of 110% value upto the delivery of site Shall be arranged by supplier and cost of the same should be included in the price quoted.

Note: Please indicate the quotation reference no. (given at the top of page 1 of this letter) on the quotation envelope.

The details of various medical equipment are also made available at www.hsccltd.co.in, modification/amendments etc, if any, shall only be notified on website.

General Manager (Procurement)

Annexure-A

Price bid

Sl No.	Name of Item	Qty(Nos)	Unit Price (Rs.) (Incl. Of forwarding, packing, insurance, transportation up to destination)	Tax (if any)	Unit Price (Incl. Tax)	Total Price(Incl. Tax) Rs.

We accept all terms and conditions of the above Invitation of Quotation

Signature

Authorized signatory Name & Designation

Technical Specification

SPECIFICATIONS OF OPERATING THEATRE LIGHT: MOBILE

1.1	Mobile operating light is required for illuminating the operating field in an emergency environment and the system can be moved from place to place.
2.1	State of the art system with shadow less light
3.1	a. Mobile light on lockable castors b. Should be LED based microprocessor control technology c. Light output 1,00,000 Lux or more d. Colour temperature 4500K or better e. Colour Rendering Index (CRI) 95 % f. Sterilizable focusing handle g. Should withstand wide voltage fluctuation h. Should have intensity control from 40-100%
3.2	Emergency Power Unit having in-built CVT with automatic change over from Mains to Battery mode in the event of power failure to provide 60 minutes back up
7.1	Should be US FDA or European CE
7.2	Manufacturer should be ISO certified for quality standards.
7.3	Electrical safety conforms to standards for electrical safety IEC 60601-1 General Requirements (or equivalent BIS Standard
7.4	Shall meet internationally recognised standard for Electro Magnetic Compatibility (EMC) for electromedical equipment: IEC-60601-1-2 :latest edition Or Equivalent BIS) or should comply with 89/366/EEC; EMC-directive as amended
7.5	Certified to be compliant with IEC 60601-2-41: Particular requirements for the safety of Operation Theatre Light or equivalent if applicable