

INVITATION FOR QUOTATION

HSCC/PUR/KCGMC/Low Value/Medical College Items/2017/MC-Dental

Dated: 13th Sept, 2017

To,

All Bidders

Subject: Invitation for Quotations for supply of Low Value Items for below mentioned dept.
1. Dept. of Dental for *Kalpana Chawla* Medical College, Karnal, Haryana.

Dear Sirs,

1. Director General, Medical Education & Research, Panchkula, Govt. of Haryana through its Consultant HSCC (India) Ltd. invites most competitive quotation for the following goods of the respective Departments as detailed below:

Annexure - (Including name of items, Specification and Quantity)

(I) Dept. of Dental

2. Quotation:
 - 2.1 The contract shall be for the full quantity as described above.
 - 2.2 Corrections, if any, shall be made by crossing out, initialling, dating and rewriting.
 - 2.3 The prices quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
 - 2.4 The unit price/ rate of the item should be clearly indicated in the tender. Rates /Prices quoted shall be Inclusive of all taxes, duties, forwarding, and insurance & transportation up to the destination, Kalpana Chawla Medical College – Karnal. Annexure-A is enclosed for price bid quotation format.
3. Each bidder shall submit only one quotation.
4. Evaluation of Quotations:

The Purchaser shall evaluate and compare the quotations determined to be substantially responsive i.e. which;

 - 4.1 are properly signed; and
 - 4.2 confirm to the terms and conditions, and specifications.
 - 4.3 final considerations of equipments shall be based on the quality of equipments during demonstration / inspection.
5. The Quotations would be evaluated item wise.
6. Award of contract:
 - 6.1 The Purchaser will award the contract to the bidder whose quotation has been determined to

be substantially responsive and who has offered the lowest rate for the item subject to quality of the items during demonstration / inspection.

6.2 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and cancel the bidding process and reject all quotations at any time prior to the issue of Purchase Order, without assigning any reason.

7. Delivery period shall be within **07** days from date of placement of order.

8. Payment shall be made only in Indian Rupees as follows:

Satisfactory Acceptance and delivery - 100% of total cost

100% Payment on submission of following documents (Duly signed & stamped at your end):-

- Copy of Purchase order.
- Consignee Acceptance Certificate in original issued by KCGMC/HSCC
- Invoice in favour of consignee/Director, KCGMC, Karnal through HSCC (I) Ltd
- Warranty Certificate in original.

9. All supplied items shall be under warranty of **12** months from the date of delivery/successful acceptance.

10. You are requested to provide your offer as follows:

Sr. No.	DESCRIPTION	Closing date & time for submission & receipt of tender	Date and time of Opening of Techno – Commercial Tenders
I.	Department of Dental	20 th September 2017 at 14:00	20 th September 2017 at 14:30

NOTE: (Under any unforeseen circumstances if the due date for submission of Tender is declared as holiday then the tender shall be submitted & opened on the next working day at the scheduled time).

The quotations will be opened in HSCC office, NOIDA as mentioned above in the presence of tenderer or their authorised representative as they may choose to attend.

11. Information brochures/ Product catalogue, if any, must be accompanied with the quotation clearly indicating the model quoted for.
12. Sealed quotation to be submitted/ delivered at the address mentioned below:

**General Manager (Procurement)
HSCC (India) Ltd.
E- 6 (A), Sector -1.
NOIDA – 201 301.**

THE COVER SHOULD BE SUPERSCRIBED WITH THE FOLLOWING:

Reference to letter of enquiry.
Due date of opening.
Serial No of the Items with Deaprtment.

13. Quoted amount should be in Indian Rupees only.

14. For all items, the Technical Evaluation Committee may opt for Demonstration of the items. The Committee may also ask for Demonstration / Inspection before supply / delivery of the items for quality assurance.

15. Insurance shall be arranged/borne by supplier.

Note: Please indicate the quotation reference no. (Given at the top of page 1 of this letter) and Serial No of the Items on the top of the envelope.

The details of various medical equipment shall be also made available at www.hsccltd.co.in, and modification/amendments etc, if any, shall only be notified on website only.

We look forward to receiving your quotation and thank you for your interest in this project.

General Manager (Procurement)

Price bid

Sl No.	Name of Item	Qty(Nos)	Unit Price (Rs.) (Incl. Of forwarding, packing, insurance, transportation up to destination)	Tax (if any)	Unit Price (Incl. Tax)	Total Price(Incl. Tax) Rs.

We accept all terms and conditions of the above Invitation of Quotation

Authorized signatory
Name , designation and sealed

DEPARTMENT OF DENTISTRY

DEPARTMENT OF DENTISTRY

Sr No.	Name of Equipment with Specification	Specifications	Requirement
1.	Mouth Mirror	Stainless steel rust resistant ISI, ISO Certified, FDA, CE, UL or BIS Approved	100
2.	Explorer	Stainless steel, ISI,ISO certified	50
3.	Probe	Stainless steel rust free, ISI, ISO Certified	50
4.	Tweezer	Stainless steel rust free, ISI, ISO Certified	100
5.	Kidney trays	Stainless steel rust free, ISI, ISO Certified	25
6.	Toffelmire	Stainless steel rust resistant ISI, ISO Certified FDA, CE, UL or BIS Approved	10
7.	Warwick James straight elevator	Stainless steel rust resistant ISI, ISO Certified, FDA, CE, UL or BIS Approved	7
8.	Air Rotor	Contra angle, ISI, ISO Certified, CE, 380000 rpm	5
9.	CheatlesForcep	Stainless steel , ISI, ISO Certified	5
10.	Glass slab	Rounded margins, thickness minimum 8mm	3
11.	Mortar pestle	Made of Bone China	3
12.	Cotton holder	Stainless steel , ISI, ISO Certified	3
13.	Stainless steel tray big	Stainless steel , ISI, ISO Certified	3
14.	Stainless steel tray small	Stainless steel , ISI, ISO Certified	3
15.	Stainless steel autoclavable drum- large	Capacity 20 litres	5
16.	Stainless steel drum for cotton	Small size	5
17.	Filling Instruments	6 sets	
	Spoon excavator	Stainless steel 0.5mm ISI, ISO Certified	6
	Spoon excavator	Stainless steel 1mm, ISI, ISO Certified	6
	Spoon excavator	Stainless steel 1.5mm, ISI, ISO Certified	6
	Burnisher	Stainless steel, ISI, ISO Certified	6
	Condenser	Stainless steel, ISI, ISO Certified	6
	Hollen back carver	Stainless steel, ISI, ISO Certified	6
	Diamond shape carver	Stainless steel, ISI, ISO Certified	6
	Amalgam carrier	Stainless steel, ISI, ISO Certified	6
	Gingival Marginal Trimmer (Mesial & Distal)	Stainless steel, ISI, ISO Certified	6
	Plastic Spatula	ISI, ISO Certified	6
	Teflon Coated Burnisher	ISI, ISO Certified	3
	Teflon Coated Condenser	ISI, ISO Certified	3
	Teflon Coated plastic Filling Instrument	ISI, ISO Certified	3
18.	Extraction Forceps		
	Maxillary Anterior forceps	Stainless steel rust free, ISI, ISO Certified	6
	Maxillary Premolar Forceps	Stainless steel rust free, ISI, ISO Certified	6
	Maxillary Molar Forceps (Right & Left)	Stainless steel rust free, ISI, ISO Certified	6
	Maxillary root forceps	Stainless steel rust free, ISI, ISO Certified	6
	Maxillary Third Molar Forceps	Stainless steel rust free, ISI, ISO Certified	6
	Mandibular Anterior forceps	Stainless steel rust free, ISI, ISO Certified	6
	Mandibular Premolar Forceps	Stainless steel rust free, ISI, ISO Certified	6
	Mandibular Molar Forceps	Stainless steel rust free, ISI, ISO Certified	6
	Mandibular root forceps	Stainless steel rust free, ISI, ISO Certified	6
	Mandibular Third Molar Forcep	Stainless steel rust free, ISI, ISO Certified	6
	Bayonet Forceps	Stainless steel rust free, ISI, ISO Certified	3
	Maxillary Cowhorn	Stainless steel rust free, ISI, ISO Certified	3
	Mandibular Cow horn	Stainless steel rust free, ISI, ISO Certified	3
	Chisel	Stainless steel rust free 2mm, ISI, ISO Certified	2
	Chisel	Stainless steel rust free 3mm, ISI, ISO Certified	2
	Periosteal Elevator	Stainless steel rust free, ISI, ISO Certified	15
	Moon's Probe	Stainless steel rust free	2
	Cryer (right and left)	Stainless steel rust free, ISI, ISO Certified	4
	Couplands elevator No. 1	Stainless steel rust free, ISI, ISO Certified	7
	Upper root forceps	Stainless steel rust free, ISI, ISO Certified	4
	Lower root forceps	Stainless steel rust free, ISI, ISO Certified	2
	Needle holder	Stainless steel rust free, ISI, ISO Certified	5
	Artery forceps	Stainless steel rust free, curved, 6 inches, ISI, ISO Certified	4

	Bone currete	Stainless steel rust free, ISI, ISO Certified	3
	Tissue forcep	Stainless steel rust free toothed tissue forceps, ISI, ISO Certified	3
	Bone ronger	Stainless steel rust free, ISI, ISO Certified	3
	Mallet	Stainless steel rust free, ISI, ISO Certified	3
	BP handle	Autoclavable Bard Parker Handle size 3	5
	Tongue depressor L shape	Stainless steel tongue depressor L shape	3
	Cheek retractors	Extraoral plastic cheek retractors made of clear plastic used for intraoral photography. Autoclavable. Hands free single piece design`	5
	19. Prosthetic Instruments		6 sets
	Stainless steel stock tray Perforated Maxillary	Stainless steel 0-4	6 sets
	Stainless steel stock tray Perforated Mandibular	Stainless steel 0-4	6 sets
	Stainless steel stock tray Non- Perforated Maxillary	Stainless steel 0-4	3 sets
	Stainless steel stock tray Non- Perforated Mandibular	Stainless steel 0-4	3 sets
	Flexible rubber bowl small		6
	Flexible rubber bowl large		6
	Lacrons carver		5
	Wax spatula		5
	Mixing spatula straight	Width approx 1"	3
	Mixing spatula curved		3
	Cement spatula		6
	Wax knife	Stainless steel	1
	Mean value articulator	Stainless steel, with remounting plates and with vertical pin. Aniodised metal	2
	20. Scaling instruments		
	Curette 4R-4L	Stainless steel rust free, ISI, ISO Certified	6
	Gracy 1-14	Stainless steel rust free, ISI, ISO Certified	2
	21. Hygienic Rubber dam kit (Adult)	Kit should Contains: -Plastic Dental Dam Frame 6" -Wedjets Stabilizing Cord Pack(S) -Hygenic Template 6" -Punch -Forcep -Rubber Dam Sheets 6x6"(36) -Set of 9 Clamps-Adult	2

Equipment big size

1.	Endomotor	<ol style="list-style-type: none"> 1. RPM range should be 250 to 1200<tel:1200> 2. Torque range should be from 0.6 to 4.0Ncm 3. Should be preset for popular rotary systems 4. Should run on battery and direct AC 5. Should be able to run for at least two hour continously 6. Programs should be customizable by the doctor 7. Inbuilt auto reverse mechanism that prevents file Separation during continuous motion 	1
2.	Apex locator	<ol style="list-style-type: none"> 1. Should have more than 95% accuracy 2. Should be able to work in both moist and dry canals and in presence of pus, blood, GP solvent 3. No preadjustment should be required before use. 4. Automatic calibration 5. Compact and light weight 6. Color display to indicate apex, short of apex and crossing of apex 7. With audible beep to indicate crossing of apex 	2

		8. Supplied with at least three file holders, five contrary electrodes, one probe cord, one canal length tester.	
3.	Implant kit with 25 implants	Implant kit with standard no. of drills and supplementary drills for different diameters, drill extender, depth gauge, screw driver, autoclavable box, implant kit for all single stage and two stage implants. Dental implants compatible with the system used: Assorted size range (2.5-6 mm diameter approx., length range 6-14 mm approx.)	1
4.	Physiodispenser with handpiece	Surgical micromotor for implant placement, meeting sterility and safety standard, speed and torque control, LCD Display with programmable setting, gear reduction.	1
5.	DigitestPulp tester	ISI, ISO Certified	2
6.	Amalgamator	<ul style="list-style-type: none"> • Individual adjustment of paste consistency • Non Capsule type • Electronically controlled • Digital display • FDA, CE, UL or BIS Approved 	1
7.	Lab type micromotor	Heavy duty Micro Motor Must be DC supply with control Unit .	1
8.	Acrylizer	With stainless steel heavy gauge chamber, fully leak proof with Argon welding, with tap for pouring of wax, inbuilt heater with auto off option at set temperatures,	1
9.	Dewaxing units	Outside & Inside body made of Stainless steel 304 grade body.inner size 300x250x350mm with digital display and controller cum digital timer.	1
11.	Hydraulic press	Heavy Duty Hydraulic Press with Pressure Gauge and In Built Oil tank	1