

AMENDMENT – II

INVITATION FOR QUOTATION

HSCC/PUR/KCGMC/Low Value/Medical College Items/2017/MC-Pathology Dated: 24th Oct 2017

Bid sale, submission and opening date of the following Items has been extended as per details given in Table -1:

| Sr. No. | DESCRIPTION | Closing date & time for submission & receipt of tender | Date and time of Opening of Techno – Commercial Tenders |
|---------|-------------------------|--|---|
| I. | Department of Pathology | 07.11.17 at 14:00 | 07.11.17 at 14:30 |

PATHOLOGY (New List)

| S.No. | NAME OF THE ITEMS | No. Required |
|-------|---|--------------|
| (B) | Morbid Histology and Morbid Anatomy | |
| 1 | Specimen Identification Solutions. | 1 |
| 2 | Paraffin embedding bath. (Preferably Imported) 1 Paraffin Reservoir capacity at least 3 liters 2 Temperature ranges: Paraffin Reservoir: 50 deg C - 70 deg C (\pm 2deg C) Work Surface: 50 deg C - 70 deg C (\pm 5 deg C) Tissue Holding Tank: 50 deg C - 70deg C (\pm 2 deg C) Cold Plate: - 5 deg C to -15 deg C to ambient 3 Refrigerant: Cold Plate, Cold Spot (Peltier controlled) 4 There should be a Membrane Keypad with LCD/LED to set and display operating parameters, current status, running time and alarm conditions for time and temperature. 5 Resolution of temperature display: \pm 1 deg C 6 Unit should have self test on power up and should display error codes in case of malfunction for easy maintenance and troubleshooting. Error codes should be indicative of the System failure or a single module failure. 7 Dimensions :(All dimensions variations \pm 10 % rounded off to integral value.) Height of Work Surface: 6 cm or more Cold Plate: (at least to hold 80 to 100 cassettes) 8 Receptacle for 6 forceps 9 Pre heated forceps of two types (for small and medium size tissue) 10 Drain Wax should remain in melted form System Configuration Accessories, spares and consumables 1 Prices should be quoted for each separately: Standard size Cassettes – 1000 Nos. 2 Large field Magnifying lens with cold light source 3 Stainless Steel Moulds of different sizes (Depth 9 to 12 mm) – 80 Nos. 4 Paraffin Scrapper – 3 Nos. 5 Halogen Bulb – 12 Nos. 6 Fuse – 12 Nos. Power Supply 1 Power input to be 220-240VAC, 50Hz fitted with Indian plug 2 Reset table over current breaker shall be fitted for protection 3 Suitable voltage corrector/stabilizer . 11.Should be USFDA and CE Europe certified. | 2 |
| 3 | Heated Paraffin Embedding Module. | 2 |
| 4 | Cold Plate for Modular Tissue Embedding System. | 1 |

| | | |
|----|--|----------|
| 30 | Plastic appointment stand for museum | 250 |
| 36 | <p>Syringes Disposable – 20cc, 10cc, 5cc and 2cc The disposable Syringes with each set having various sizes ranging from 2cc to 20cc.</p> <p>20cc----- 10 set 10cc----- 75 set 5cc-----10 set 2cc----- 5 set</p> | 100 sets |
| 39 | <p>LED TV with DVD Player with USB Slot 40 INCHES along with DVD Player</p> <ol style="list-style-type: none"> 1. USB 2 Side Support. 2. Fully HD & HDMI Support. 3. AV/Composite input slot compatible with newer devices. 4. Built in Woofer & Wi-Fi 5. Photo Sharing Plus 6. Photo Frame Mode 7. USB - Super Multi Format Play09oi 8. Mobile High-Definition link (MHL) 9. FM Radio 10. Solid Durability With Enhanced Protection Features. <p>Power: AC 100-240 V 50/60 Hz Other: Five Star Rating</p> | 1 |

General Manager (Proc.)

