

HSCC (India) Ltd

AMENDMENT No – V Dated 01.10.2014

Sub.: Procurement of Medical Equipment for NEIAH, Shillong.

Ref: Tender Enquiry No.: HSCC/PUR/NEIAH/Medical Equipment/2014 dated 19.8.2014

ITEM NO 1: Digital X-Ray Machine (1000 MA)

Existing Tender Specification	Amended/Read As
Out of X-Ray generator, X-Ray Tube & detector, two should be of the same make/manufacturer and specify their manufacturer's name.	Deleted
Please mention tube loading for small focus and large focus, should be at least 30KW or more for small focus and at least 80KW for large focus.	Please mention tube loading for small focus and large focus, should be at least 30KW or more for small focus and at least 65KW for large focus
Dry Chemistry Camera. Should have minimum 500 DPI or more and should print at least 3 sizes of films on line out of 10x12,10x14,11x14, 8x10 and 14x17 inches.	Dry Chemistry Laser/Thermal Camera. Should have minimum 500 DPI or more and should print at least 3 sizes of films on line out of 10x12,10x14,11x14, 8x10 and 14x17 inches

ITEM NO 7: Electric Hydraulic Ophthalmology Operation Table 1 No.

Existing Tender Specification	Amended/Read As
4 System Configuration Accessories, spares and consumables e. Lithotomy leg holders 'Geopel type' (Adult and Paediatric) – 1 set each h. Accessories for operating in prone position i. Kidney bridge	4 System Configuration Accessories, spares and consumables e. Deleted h. Deleted i. Deleted

ITEM NO 9: SHADOWLESS LAMP (CEILING-SINGLE DOME)

Existing Tender Specification	Amended/Read As
Should have intensity control from 40-100%	Should have max. light intensity control variation

HSCC (India) Ltd

ITEM NO 12: VIDEO COLONOSCOPY SYSTEM

Existing Tender Specification	Amended/Read As
Should be fully upgradable with EUS and balloon enteroscopy.	Deleted
Light source - Combined or separate 100 watt or more Xenon with emergency backup facility.	Light source - Combined or separate 150 watts or more Halogen/ 100 watt or more Xenon with emergency backup facility.

ITEM NO 14: SHADOWLESS LAMP (MOBILE-SINGLE DOME)

Existing Tender Specification	Amended/Read As
Should have intensity control from 40-100%	Should have max. light intensity control variation

ITEM NO 16: ELECTROCAUTERY/SURGICAL DIATHERMY WITH ACCESSORIES

Existing Tender Specification	Amended/Read As
Accessories : Point to be Added: Standard set of all purpose monopolar & Bipolar cutting & coagulation electrodes	<u>Accessories</u> : Standard set of all purpose monopolar & Bipolar cutting & coagulation electrodes Acc. Shall broadly cover the following: 4.1 Main Unit -01 No 4.2 Pencil Holder -01 No 4.3 Electrode Blades -10 Nos 4.4 Electrode Needle Fine -10 Nos 4.5 Electrode – Ball point – (3 different sizes- small, Medium & long) – 5 each 4.6 Pencil Switches with hand control -10 Nos disposable, 2 Nos reusable 4.7 2 Foot Control double paddle 4.8 chuck handles - 10 Nos disposable, 2 Nos reusable 4.9 Reusable Patient plate/Neutral Electrode with cable for

HSCC (India) Ltd

	<p>Adult & Children – 2 each. Disposable patient plate with cable for Adult & Children – 100Nos. each</p> <p>4.10 Mobile trolley/cart on castors with brakes -01 Nos</p> <p>4.11 Power cord of 5 meters length and plug with appropriate rating for standard Indian electrical sockets. -01 No.</p> <p>4.12 Reusable & sterilizable (5nos.) & disposable electrode handle with finger switch</p> <p>4.13 Set of electrodes (flat tip short, flat tip long & pin point- 5nos. each) with electrode container with holder</p> <p>4.14 Tip cleaner</p> <p>4.14 Bipolar forceps with cable</p> <p>4.15 The system should be capable of accepting standard accessories of major international brands.</p> <p>In case missed out, bidders should supply all standard set of accessories with the equipment for monopolar & Bipolar requirements as per the specifications.</p>
<p>Monopolar cutting- 250 W or more, Monopolar coagulation- 100 W or more, Bipolar cutting- 120 W or more, Bipolar coagulation- 100 W or more</p>	<p>Monopolar cutting- 300 W or more, Monopolar coagulation- 100 W or more, Bipolar cutting- 130 W or more, Bipolar coagulation- 100 W or more</p>

ITEM NO 23: Bone Drill Machine & Bone Cutter

Existing Tender Specification	Amended/Read As
The drilling speed should be more than 900 RPM and torque should be available upto 11Nm.	The drilling speed should be more than 900 RPM and torque should be available atleast 11Nm or equivalent in/lb. Reaming speed should be 200 rpm or more. Drill should be lubrication free.
The System should have dedicated oscillating saw hand piece.	The System should have dedicated oscillating/ sagittal saw hand piece.
The oscillating saw handpiece should have rotating head at 4 intervals in 90deg each	The oscillating saw handpiece should have rotating head at 4 intervals in 90deg each/ at 8 intervals in 45deg each
	The complete System should be from one manufacturer.

Commercial Amendments:

Existing Clause

Section – IX Qualification Criteria (Page No. 123)

01. The Tenderer must be a Manufacturer. In case the manufacturer does not quote they may authorise their agent as per proforma of Manufacturer authorization form as given in the tender enquiry document to quote and enter into a contractual obligation.
02. (a) The Manufacturer should have supplied and installed in last **Five** years from the date of Tender Opening, atleast 50% of the quoted quantity of the similar equipment meeting major specification parameters which is functioning satisfactorily. The foreign Manufacturer satisfying the above criteria should also have supplied and installed in last **Five** years from the date of Tender Opening, at least 50% of quoted quantity of similar model which is functioning satisfactorily any where outside the country of manufacture.
02. (b) The Tenderers quoting as authorized representative of the manufacturer meeting the above criteria 02 (a) should have supplied and installed in last **Five** years from the date of Tender Opening, atleast 50% of the quoted quantity of similar equipments which is functioning satisfactorily, any where in India of any manufacturer.

To be Read as:

1. The Tenderer shall be a Manufacturer or its authorised agent.
- 2(a)The Manufacturer should have supplied and installed in last **Five** years from the date of Tender Opening, atleast **50%** of the quoted quantity (subject to minimum 1 nos.) of the similar equipment meeting major specification parameters which is functioning satisfactorily. The foreign Manufacturer satisfying the above criteria should also have supplied and installed in last **Five** years from the date of Tender Opening, at least 50% of quoted quantity (subject to minimum 1 nos.) of similar model which is functioning satisfactorily any where outside the country of manufacture.
- 2.(b))The Tenderers quoting as authorized representative of the manufacturer meeting the above criteria 02 (a) should have supplied and installed in last **Five** years from the date of Tender Opening, atleast **25%** of the quoted quantity (subject to minimum 1 nos.) of similar equipments which is functioning satisfactorily, any where in India of any manufacturer.

HSCC (India) Ltd

In view of the above Amendments, Bid submission date is hereby extended from 09.10.2014 to 16.10.2014. The validity of bid security (EMD) shall be considered from the original date of submission of bid i.e. from 19/9/2014.

Bidders should submit all the documents as mentioned in the Check List, GIT Clause 11, Section-IX and elsewhere mentioned in the Bid Document.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPMP website for corrigendum /amendments etc. if any, as these will be notified on these portal only. No separate advertisement will published in the news papers in this regard.

Director, NEIAH Shillong