

AMENDMENT No. 1 dated 15.07.2014

Ref: Tender Enquiry No.: HSCC/PUR/AIIA/Panchakarma /2014 dated 16/06/2014

A. All Bidders are requested to note the following clarification/ amendments after the pre bid meeting held on 25.06.2014.

Sr.	Existing clause	Amendment requested	Clarification/ Amended as
1.	The estimated value Rs. 1.06 Cr. (approx.) (Clause 11 on Page no 6)	(i) Bidder requested to amend the lowest estimated cost around Rs. 2.12 Crore due to change in technical specification. (ii) Bidder requested to amend the lowest estimated cost around Rs. 1.85 to 1.96 Crore due to change in technical specification.	Estimated cost Rs. 1.85 Crore and EMD amount Rs.3,70,000/-
2.		Bidder requested to amend the following in technical specification:	
3.	Droni (Item no. 1): <i>a. Wooden</i> Made out of single piece Medicinal wood such as..... <i>b. FRP/FIBERGLASS</i> Heavy Duty Droni made of NPG- ISO and glass fiber...	Suggest Teak wood & mill jointed wooden board. Suggest FRP to be changed to Teak wood.	As per tender
4.	Steam Bath system (Item No. 2): 2(a) Steam Bath Chamber sitting Made of FRP moulded with heat resistant ISO resin and	Suggest FRP to be changed to Teak wood.	As per Tender
5.	Nadi Swedan Yantra (Item no. 3): Should be designed to work with only 6 ltrs of water to save...	Suggest 12 ltrs water capacity to have 4-5 steam baths continuously.	As per Tender
6.	Shirodhara system (Item no. 4) <i>c) Shirodhara Patra</i> FRP Vessel made of heat.....	Suggest Stainless Steel Vessel for longer working life.	As per Tender
7.	Awagha Tub (Item no. 7): Full body Fibreglass bath tub	Suggest Stainless Steel sheet for heavy work load & longer working life of unit and is more hygienic.	As per Tender
8.	Virechan Peeth (Item no.11): Seat made of wood finish NPG- ISO Composite	Suggest seat to be made of Wood.	As per Tender

B. All Bidder are also requested to note the following amendments in the tender enquiry document sale, submission and opening date of aforesaid tender:

Sl. No.	Description	Revised Schedule
i.	Last dates and time of sale of tender enquiry documents.	31.07.2014, 10:00 hrs to 17:30 hrs IST
ii.	Last date and time for submission of bid.	01.08.2014, 14:30 hrs IST
iii.	Time and date of opening of Techno – Commercial bids.	01.08.2014,15:00 hrs IST

All other terms and conditions of the tender enquiry document shall remain unchanged.

s/d
Director (AIIA)