

All Bidders**Amendment -I**

Subject: Supply, Installation, Testing and Commissioning of Modular Operation Theatres at AIIA, Sarita Vihar, New Delhi.

IFB No. : HSCC/SES/AIIA/MOT/2013

Dear Sir,

This has reference to above IFB No. for the Subject works.

The following Amendment may be noted which shall be treated as part of the contract to be submitted duly signed & stamp along with tender.

	Bidder's Query	Amendment/Clarification
1	<p>M/s MPS Clause 2.2 (ii), PQ</p> <p>Experience of having successfully completed similar work during last 7 years ending last day of month previous to the one in which tenders are invited should be either of the following :</p> <p>Three similar* completed works costing not less than the amount equal to 40% of the estimated cost. or</p> <p>Two similar* completed works costing not less than the amount equal to 50% of the estimated cost. or</p> <p>One similar* completed work costing not less than the amount equal to 80% of the estimated cost.</p> <p>One completed work of any nature (either part of 2.2,(ii) or separate one costing not less than the amount equal to 40% to the estimated cost with some Central/State Government organisation/Central Autonomous body/Central Public Sector Undertaking.</p> <p>*Similar nature of works means successful completion of Supply, Installation of Medical Gas Manifold System.</p> <p>A Certificate from client for completion of work(s) must be submitted along with application. Own works/ Certification of agencies shall not be considered for prequalification.</p>	<p>Tender Terms & Conditions Prevail.</p>

	<p>It is suggested that The TDS certificate should be asked to be submitted to justify the order value against experience of similar nature of work in case of the project executed in private sector.</p> <p>Because without submission of TDS certificate the value of work done in private sector can not be verified.</p>	
2	<p>Vol-III, Clause No. 1</p> <p>Three Technologies of material ie. PUF, Stainless Steel & SMCS is specified.</p> <p>We understand that PUF is locally available. Stainless Steel/SMCS is imported as per the specifications mentioned in the tender. Hence there is no price comparison between Indian and Imported material.</p> <p>SMCS being in built antibacterial properties, scratch less, dent free, washable and long lasting which calculates to low cost over a span of period. It is not clarified whether material of wall cladding should be imported or Indian. As per our experience and specifications of the tender the material should be imported. We suggest you for imported make. If Indian make material is recommended then on site antibacterial paint should be allowed.</p>	Tender Terms & Conditions Prevail.
3	<p>Vol-III, Clause No. 10(a)</p> <p>Boom Suspension for 26" HB Progressive Scan Flat Panel Monitor.</p> <p>Since OT Light with Camera & Monitor (Imported) is mentioned at Clause No. 21. This specification seems to be part of OT Light. Hence should be deleted from Equipment Boom System & be mentioned with OT Light OR be clarified.</p>	<p><u>Item -Boom Suspension for 26" HD Progressive Scan Flat Panel Monitor -Deleted</u></p> <p><u>Description</u> : Should be 26" High Definition Progressive Scan Flat-panel Monitors with ceiling mounted spring arm suspension in tandem-mount to support high-definition/HDTV progressive Scan images and should be able to support and display DVI/HDTV, RGBHV, S-Video, Composite video signals. - Deleted</p> <p>The flat Panel suspension should be ready with the cables for integration of High Definition Digital (DVI/HDTV), RGBHV(High Resolution), SVHS(S-Video), Composite video signals to travel from the various sources of video like endoscopic camera, room camera, in light camera, high definition flat panel monitors, while assuring native resolution / signal. -Deleted</p>

		Such 26” Flat-panel High Definition Progressive Scan Monitor with spring-arm suspension should meet all criteria mentioned above. – Deleted
4.	<u>General</u> We further suggest that only established brands of equipments should be procured. Merely meeting the technical parameter new make of products should not be procured.	Tender Terms & Conditions Prevail.
5	Clause 10.0 of Technical Specification ADJUSTABLE MOVABLE BOOM ARM SYSTEMS –Imported a) Equipment Boom System with boom suspension for Progressive Scan Flat Panel	Clause 10.0 of Technical Specification may be read as ADJUSTABLE MOVABLE BOOM ARM SYSTEMS – Imported a) May be read as Equipment Boom System.
6	Item 2.0 of Tender Specification Ceiling Filtration System/Laminar Flow System -Laminar Air Flow system should comply with DIN 1946. -Complete Air Management System should be duly CE Certified. Item 1.0 of Tender Specification Walls & Ceiling Construction The wall panel should be fixed to the brick wall....., no welding should be allowed.	Item 2.0 of Tender Specification Ceiling Filtration System/Laminar Flow System. <u>-Laminar Air Flow system should comply with DIN-1946-</u> Deleted. -Complete Air Management System should be duly CE Certified –CE Certified deleted. Item 1.0 of Tender Specification Walls & Ceiling Construction Word No welding should be allowed- Deleted
7.	Item No. 4.0 of Tender Specification Door & Frames (Hermetically Sealed Doors) -Controller should be Microprocessor based Controller (CE Marked) and should have digital display.	Item No. 4.0 of Tender Specification Door & Frames (Hermetically Sealed Doors) -Controller should be Microprocessor based Controller (CE Marked) and should have digital display- CE marked and digital display deleted.
8	Item No. 5.0 of Tender Specification The touch screen typed Control Panel should be 19” medical grade color TFT/LED panels with 1280 x 1024 (SXGA) resolution stationed in sterile field.	1280 x 1024 (SXGA)- Deleted
9	Item No. 21 of Technical Specification OT Light with Camera and Monitor (Imported) Rotation -360 degree	Item No. 21 of Technical Specification OT Light with Camera and Monitor (Imported) Rotation may be read as 330 degree.

Bidder should follow the tender terms & condition for the unanswered queries.

All other terms & conditions remain unchanged.

CGM (D&E & Projects)