

AMENDMENT – I

Dated: 28.02.2014

Tender Enquiry No.: HSCC/PUR/AIIMS/RAEBARELI/ TEMP OPD/2014 Date: 24.02.2014

LIMITED TENDER ENQUIRY

Sub: Procurement of Medical Equipment for setting up of Temporary OPD for AIIMS, Rae Bareli
Under PMSSY

TO BE ADDED:-

	Gyn. OPD - 2 Nos.	Qty.
1	Led spot examination light.	2
	Ophthalmology - 2 No.	Qty.
2	Lensometer	2

SPECIFICATIONS:

LED SPOT EXAMINATION LIGHT

Should be with LED Examination light, mounted on a pedestal stand Lux Output 50000

LED lamp with 50000 to 100000 hours lamp life

Cool Beam at similar to that of the day 5500 deg K.

Illumination field 30cm diameter or more.

Arm Flexibility for easy positioning.

Should be European CE / US FDA standard equipment.

LENSOMETER / FOCIMETER

- For Accurate analysis of the power of any lens including multifocal, prisms, contact lens and centration.
- Specifications
- Light source: LED/ Halogen bulb
- Sphere measuring range: -25 +25D unit 0.25D
- Axis Measuring range 0-180° units 1°
- Prism Measuring range 0-6 unit 1 Prism diopter
- Scale reading : internal -/+ external
- Lens diameter: 24-90 mm
- Eyepiece Focusing: +7D to -10D
- Tilt angle: 0-90° , freely adjustable
- Power supply: AC 220 to 240 volts as well as battery operated
- Accessories: 1 Dust cover, 1 contact lens holder

Should have European CE or US FDA approved.

All other terms and conditions of the tender enquiry remain unaltered