

AMENDMENT No - III

Dated 16.06.2016

**Subject: Amendment to the tender Enquiry Document**

**Ref: Tender Enquiry No.: HSCC/SJH/Med. Eqpt./18 dated 18.05.2016**

Due to inadequate response, the bid submission date for Item No. 2, 3, 8, 10, 11, 14, 16, 17 is extended to 27.06.2016.

**TECHNICAL SPECIFICATIONS**

**ITEM No. 8**

**SPECIFICATION FOR OPERATING LAPAROSCOPE WITH CO2 INSUFFLATOR, FULL HD CAMERA SYSTEM & ACCESSORIES (1 set)**

**Existing as:**

System should have facility to offer various visualization modes for surgery and diagnosis by shifting the color spectrum like **BLUE & GREEN** light for recognition of the finest tissue Structures and their differentiation

**To be read as:**

**To be quoted as optional:**

System should have facility to offer various visualization modes for surgery and diagnosis by shifting the color spectrum like **BLUE & GREEN** light for recognition of the finest tissue Structures and their differentiation.

➤ **Existing as:**

**Xenon Light Source:**

Xenon light source of 300 Watts

➤ **To be read as:**

Point Deleted :- (Xenon Light Source)

➤ **Existing as:**

Or alternatively

**LED Light source:**

LED light source of 170 Watts or more

➤ **To be read as:**

**LED Light source:**

**Amended as:** - LED Light source 300 watts or more.

**Grasping Forcep Fenestrated:**

**Amended as:** -Atraumatic serration, fenestrated, grasping forcep with unipolar connection, insulated sheath, handle without ratchet. Should be dismantable into three parts namely, outer tube, handle and inserts. Length **between 33-36cm**, size: 5.0mm. – 04 Nos.

**Grasping Forceps Fenestrated curved:**

**Amended as:** Atraumatic, fenestrated, curved grasping forcep with unipolar connection, insulated sheath, handle without ratchet. Should be dismantable into three parts namely, outer tube, handle and inserts. **Length between 33-36cm**, size: 5.0mm – 04 Nos.

**Grasping Forcep Atraumatic:**

**Amended as:** -Atraumatic, fine serration grasping forcep with unipolar connection, insulated sheath, handle without ratchet. Should be dismantable into three parts namely, outer tube, handle and inserts. Length between **33-36cm**, size: 5.0mm – 04 Nos.,

**Bowel Grasping Forcep:**

**Amended as:** Bowel Grasper fenestrated, size: 5.0mm, length **33-36cm**, handle with ratchet, insulated shaft.Dismountable into three parts. – 02 Nos.

**Babcock Grasping Forcep:**

**Amended as:** Babcock Grasper atraumatic, size: 5.0mm, length **33-36cm**, handle with ratchet, insulated shaft.Dismountable into three parts – 02 Nos.

**Allis Grasper:**

**Amended as:** Allis Grasper, size: 5.0mm, length **33-36cm**, handle with ratchet, insulated shaft.Dismountable into three parts. – 02 Nos.

**Unipolar Tooth Grasping forcep:**

**Amended as:** Tiger Jaw 2x4 teeth insulated. Dismantable into different parts, insert, handle and working tube, Working length of **33-36 cm**, size: 5mm. Handle with ratchet. -02 Nos.

**Scissor curved Unipolar:**

**Amended as:** METZENBAUM curved scissor, length of blade 12mm, connection for unipolar HF cable, dismountable into insert, tube and handle. Working **length 33-36 cm**, size: 5mm, handle without ratchet. – 02 Nos.

Similar Change may be considered for Scissor curved unipolar, curved dissecting & Grasping forceps & other for cops & graspers.

All other terms & conditions remains unaltered.

**Medical Superintendent  
Safdarjung Hospital & VMMC,  
New Delhi.**