

HSCC/PUR/LHMC/2014

Date: 22.9.2014

**Amendment-I**

Ref.: Tender Enquiry No. HSCC/PUR/LHMC/2014 dt. 16.7.2014.

Sub.: Procurement of Radiotherapy Equipment for Lady Hardinge Medical College & Hospital, New Delhi – **(For Schedule 5 Isocentric Mobile C-Arm X-Ray Imaging System and Schedule 6 Dosimetry, Radiation Safety & Quality Assurance Systems/Tools).**

**The revised date of submission of bids on e-tender basis is changed from 23<sup>rd</sup> Sept. 2014 to 08<sup>th</sup> Oct. 2014.**

Bidders must submit Technical Bid as per Amendment-III dt. 22.8.2014 (w.r.t Schedule 1 to 4) and upload & submit all other documents as mentioned in Page No. 8 & 9 of the Bid Document.

Tender Fee of Rs.5000/- is one-time fee. Bidder can quote for any or all equipment on submission of Tender Fee. Bidder should submit original Tender Fee & original EMD as per Bid Document in their Technical Bid refer Amendment – III dt. 22.8.2014 (w.r.t Schedule 1 to 4) and must be submitted on or before due date and time.

In case, quoting for more than one Schedule/Equipment, original EMD as per Bid Document & photocopy of Tender Document Fee (as proof) must be submitted along with Technical Bid and must be submitted on or before due date and time.

As it is e-tender, tender format of “**Tender Form with Price**” is to be uploaded along with Price Schedules. Further, tender format of “**Price Schedule in Foreign Currency**” is amended without any change but only allowing the bidders to type/quote the prices in words and figures in the cells/field which are blocked.

Bidders must properly check the Bid Document along with this present Amendment for the proper submission of bids on e-tender basis on or before due date and time.

Rest all tender conditions remain unchanged.

Director,  
LHMC, New Delhi

Encl.: As above.

**Amendment- I**

Ref.: Tender Enquiry No.: HSCC/PUR/LHMC/2014 Dated 16.07.2014.

Sub.: Procurement of Radiotherapy Equipment for LHMC, NewDelhi.

	<b>Isocentric Mobile C-arm X-ray Imaging System</b>	
	3.1.5The system shall have vertical C-Arm travel capability (preferably 16° higher) to adjust the imaging chain height.	3.1.5The system shall have vertical C-Arm travel capability to adjust the imaging chain height.
	5.1 All approvals such as FDA, IEC etc should be enclosed for the quoted model.	5.1All approvals such as FDA or European CE should be enclosed for the quoted model.
		Bidders must visit LHMC, New Delhi hospital site and quote as per the site conditions, tender conditions and amendments.
	<b>5.Mould Room, Patient Fixation And Immobilization Accessories (Dosimetry)</b>	
	The vendor shall provide 4 complete sets of Universal All in One Multipurpose Base Plate, upgradable to SBRT, SRS/SRT frameless, made of carbon fibre, having a total solution for adult and pediatric to treat all body sites viz. Head, Head & Neck, Breast, Thorax, Abdomen, Pelvic SBRT, SRS & SRT. Ct Markers – 300 nos.	The vendor shall provide 4 complete sets of multipurpose Base Plate, upgradable to SBRT, SRS/SRT frameless, made of carbon fibre, having a total solution for adult and pediatric to treat all body sites viz. Head, Head & Neck, Breast, Thorax, Abdomen, Pelvic SBRT, SRS & SRT. Ct Markers – 300 nos.
	<b>Commercial Amendment</b>	
	<b>GIT Clause 34. Comparison of Tenders</b> Net Present value (NPV) of the Comprehensive Annual Maintenance	<b>GIT Clause 34. Comparison of Tenders</b> Net Present value (NPV) of the Comprehensive Annual

charges (CMC) quoted for 3 years after the warranty period shall be added to the bid price for evaluation and will be calculated after discounting the quoted price by a discounting factor of 10% per annum.”	Maintenance charges (CMC) quoted for <b>5 years</b> after the warranty period shall be added to the bid price for evaluation and will be calculated after discounting the quoted price by a discounting factor of 10% per annum.”
--	---

**The revised due date of submission of bids on e-tender basis is changed from from 23.9.2014 to 08.10. 2014.**

Bidders shall be entirely responsible for complete installation, testing & commissioning of the equipment, in case any items are not mentioned inadvertently in the Bid Document including Amendments.

Rest all remains unchanged as per the Bid Document.

**Director,  
LHMC, New Delhi**