

Amendment no I Dated 20.02.2017
IFB no. HSCC/KCGMC/Medical Equipment/2016/off/C/01/F Dated 03.02.2017

All bidders may please note the following:

Item no.	Tender technical Specifications	May please read as
Auto pipettes for sr. 1 to 13	Should have secure volume lock to deliver accurate volume and a control button with very low operating force. Colour indication for pipette volume.	Should have secure volume lock to deliver accurate volume and a control button with very low operating force.
	Should have a '4 Digits with magnifier' Volume Display	Should have a 3 or 4 Digits Volume Display.
	To provide Thermal, Mechanical and Chemical Stability the Piston should be manufactured of non-corrosive, non-reactive light polymer material like a combination of Fortron and PEEK material.	To provide Thermal, Mechanical and Chemical Stability the Piston should be manufactured of non-corrosive, non-reactive material.
	Should be Autoclavable.	Should be Autoclavable /Lower part of Pipettes should be Autoclavable
	Micropipettes 0.1 to 10ml. Variability 1 ml increment	Micropipettes 1 to 10ml. Variability 1ml increment.

As per SCC clause no. 9, the validity of bid security (EMD) shall be considered from the original date of submission of bid/ tender in case tender opening date extended.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

s/d

CGM, HSCC (I) Ltd.
For & on behalf of DGMER, Panchkula