

All Bidders**Amendment -VII**

Subject: Supply, Installation, Testing and Commissioning of Integration and Data Management System for Modular Operation Theatres for Jai Prakash Narayan AIIMS Trauma Centre, New Delhi.

IFB No. : HSCC/SES/IOT/AIIMS-TRAUMA/2015

Dear Sir,

This has reference to above IFB No. for the Subject works.

The following Amendment may be noted which shall be treated as part of the contract to be submitted duly signed & stamp along with tender.

Sr. No.	Bidder's Query				Clarification/Amendment
	Point 2.2-10x6 Matrix is available with single manufacturer				Point 2.2- The audio/video router should have the following outputs.
	<u>As per the last Amendment-VI</u>				
	S. No	I/P (Source)	Signal Type	Output (Destination)	Matrix size should be 12/6 or 10/6 For 12/6, One for Archiving system and another for Free for Future use.
	1.	Endoscopic Camera	DVI/HD-SDI	DVI Monitor - 1	
	2.	In-Light Camera	S-Video/DVI	Monitor -2	
	3.	Room Camera	VGA/DVI	Conference Room	
	4.	C-Arm	DVI	Monitor – 3	
	5.	Navigation System	VGA	Archiving System	
	6.	Microscope	DVI	Monitor -I	
	7.	Patient vital sign	DVI/VGA		
	8.	Archiving System	DVI/VGA		
	9.	Free for Future Use	DVI		
	10.	Conferencing Solution	DVI/S-Video		

	(Polycom)			
--	-----------	--	--	--

The bid submission date is extended from 26.05.2016 to 02.06.2016 and bid security should be valid for 180 days from the date of bid submission ie. from 02.06.2016.

All other terms & conditions remain unchanged.

Chief General Manager
For & on behalf of Director (AIIMS)