

HSCC (India) Ltd

AMENDMENT – VII

Dated: - 29.04.2016

Subject: Amendment to the tender Enquiry Document.

Ref: Tender Enquiry No.: HSCC/KCGMC/ Medical Equipment / 2015-16/02 Dated 25.02.2016

Bid sale, submission and opening date for item no. 1,3 & 5 has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	11.05.2016, 2.00 P.M
ii.	Closing date & time for receipt of tender	11.05.2016, 2.30 P.M.
iii.	Time and date of Opening of Techno – Commercial Tenders	11.05.2016, 3.00 P.M.

Bidders are requested to note the following technical Amendment:

Item No. 01 - a). Diagnostic Cystoscope, b) Cystoscope - Paediatrics, c) Resectoscope

S.No.	Description	Amendment by HSCC/ KCGMC
	Technical Specification:-	
Part B	High Definition Camera System for Resectoscope	
	Each set should consists of following items:	
1	Full High Definition Digital Camera	
•	The system should have Optical zoom to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used having focal length anything between 15 - 31 mm (2X) or better. Camera & Coupler should be one piece (Integrated)	The system should have Optical zoom to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used having focal length anything between 15 - 31 mm (2X) or better. Camera & Coupler should be one piece for better functioning to avoid seepage of solution & damage to camera while sterilization.

HSCC (India) Ltd

Item no. 3 Pulse Oximeter, Qty : 17

a	<u>Description of Function:</u>	Amendment by HSCC/ KCGMC
g	Standard, Safety & Training:	
	Should be US FDA or European CE	Should be US FDA & European CE

Item No. 5

Name of Item : Defibrillator with recorder & Monitor.

a	<u>Description of Function:</u>	Amendment by HSCC/ KCGMC
g	Standard, Safety & Training:	
6.	It should have ability to measure chest compression rate & depth in real time with both visual & audible feedback & optional CPR index on screen	Stands deleted
9.	It should have ability to filter out CPR artifacts & allowing person to see organized rhythms without interrupting chest compression.	Stands deleted
	The Should have facility to monitor following parameters : a. Spo2 b. Etco2 c. NIBP d. ECG	The defibrillator should be facility to monitor following : a. Masimo Spo2 suitable for all perfusion patients, capnostat, Mainstream ETco2, NIBP & ECG
13.	In addition to standard accessories following items have to be supplied with unit.	
b.	NIBP Paediatric cuff with hose – 1 No.	Both paediatric & adult cuff with hose.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

s/d
CGM, HSCC India Limited
For and on behalf of DGMR, Panchkula