

## HSCC (India) Limited

Dated: 18.12.2024

### AMENDMENT-VIII

**Subject: Retender for “Planning, Detailed Design and Engineering, Construction, Testing, Commissioning & Validation of Bio Safety Level-3 laboratory at Govt. Medical College, Kota, Rajasthan including all infrastructure works in relation thereto such as Civil, Electrical, Mechanical, HVAC, Plumbing, Drainage, ETP, Furniture & Bio Safety Cabinets etc., Operation & Maintenance during one (01) year of Defect Liability Period and O&CMC for four (04) years after DLP on EPC Mode”.**

**Tender No: HSCC/SES/BSL-3/EPC/GMC-Kota/2024/72 Dated: 04.10.2024**

With reference to the above cited subject work, the following Amendment may be noted, which shall be treated as a part of the contract to be uploaded along with tender/ bid:

**1. Reply to Queries raised by bidder.**

<b>Sr. No</b>	<b>Vol./ Cl.No. / Ref.</b>	<b>As Per Tender Condition</b>	<b>Bidder Queries / request</b>	<b>HSCC Reply / to be amended</b>
1.	Vol-4 DBR & Tech. Spec.	<b>H.T. Sub-stations including VCB panel, Transformers, bus ducts, HT cables</b> etc and electrical cabling for connection to the BSL building through trenching.	Contradicting with following mentioned on Pg no-258 Electricity supply for meeting electrical load requirements of BSL-3, Kota, tapping from Existing substation of campus by the EPC Contractor.	Scope Shall be as per Electrical DBR. Sub Station not in Scope
2.	Vol-4 DBR & Tech. Spec.	Main LT, <b>Capacitor panels (APFC), AHF, Rising mains, MV Panels</b>	Please confirm.	Substation not in scope i.e. Main LT Panel, APFC, AHF, rising main are not in scope.
3.	Vol-4 DBR & Tech. Spec.	Bio-Safety Cabinets –NSF 49 compliant ( A2 type -1 Nos. for each BSL -3 & BSL-2 and one B2 type -1 Nos)	Requesting approval on EN12469 certified bio safety cabinets which is equivalent.	Biosafety cabinets- NSF 49 compliant/ EN12469 certified as per DBT guidelines.

All other terms & Conditions of the Tender shall remain unchanged.

The above amendment shall be treated as integral part of the tender document and to be submitted duly signed & stamp along with tender/bid.

Prospective bidders are advised to regularly visit through HSCC e-tender portal <https://hsc.enivida.com> & HSCC website <http://www.hsccltd.co.in> as corrigendum/amendments etc., if any, will be notified on this portal only and separate advertisement will not be made for this.

(- Sd -)  
General Manager (Procurement)  
HSCC (India) Ltd.