

All Bidders**Amendment -VIII**

Subject: Supply, Installation, Testing and Commissioning of Integration and Data Management System for Modular Operation Theatres for Jai Prakash Narayan AIIMS Trauma Centre, New Delhi.

IFB No. : HSCC/SES/IOT/AIIMS-TRAUMA/2015

Dear Sir,

This has reference to above IFB No. for the Subject works.

The following Amendment may be noted which shall be treated as part of the contract to be submitted duly signed & stamp along with tender.

Point 2.2- Matrix size should be 12/6 instead of 12/6 or 10/6 and should be as per Amendment-V as per detailed below.

As per the Amendment-V

S. No.	I/P (Source)	Signal Type	Output (Destination)
1.	Endoscopic Camera	DVI/HD-SDI	DVI Monitor -1
2.	In-Light Camera	S-Video/DVI	Monitor -2
3.	Room Camera	VGA/DVI	Conference Room
4.	C-Arm	DVI	Monitor – 3
5.	Navigation System	VGA	Archiving System
6.	Microscope	DVI	Monitor -I
7.	Patient vital sign	DVI/VGA	
8.	Archiving System-1	DVI/VGA	
9.	Free for Future Use	DVI	
10.	Free for Future Use	DVI	
11.	Archiving System-II	DVI/S- Video	
12.	Conferencing Solution (Polycom)	DVI/S- Video	

The bid submission date is extended from 02.06.2016 to 06.06.2016 and bid security should be valid for 180 days from the date of bid submission ie. from 06.06.2016.

All other terms & conditions remain unchanged.

Chief General Manager
For & on behalf of Director (AIIMS)